



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118806

Slope : 98.7

pH Meter ID : WC PH MERER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/24/2022	10:30
2	CAL2	1	Water	NA	NA	20.4	7.01	02/24/2022	10:31
3	CAL3	1	Water	NA	NA	20.4	10.03	02/24/2022	10:32
4	ICV	1	Water	NA	NA	20.5	6.99	02/24/2022	10:35
5	CCV1	1	Water	NA	NA	20.4	2.01	02/24/2022	10:40
6	N1652-01	1	Water	NA	NA	20.5	4.46	02/24/2022	10:43
7	N1652-01DUP	1	Water	NA	NA	20.6	4.47	02/24/2022	10:45
8	CCV2	1	Water	NA	NA	20.5	12.03	02/24/2022	10:46

13118806

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1652 WorkList ID : 157111 Department : Wet-Chemistry Date : 02-24-2022 10:23:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/23/2022	Water	N1652-01	pH	Cool 4 deg C	STER08	N21	POOL-WATER	02/23/2022	9040C

Date/Time 02/24/22 10:25
Raw Sample Received by: CP
Raw Sample Relinquished by: CP

Date/Time 02/24/22 11:30
Raw Sample Received by: CP
Raw Sample Relinquished by: CP