



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB118814

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP95449
Turbidity Calibration std, 0NTU	WP95448
Turbidity Calibration std, 20NTU	WP95443
Turbidity Calibration std, 40NTU	WP95442
Turbidity Calibration std, 1NTU	WP95446
Turbidity Calibration std, 5NTU	WP95445
Turbidity LOQ std, 1.0NTU	WP95450
Turbidity Calibration - CCV std, 10 NTU	WP95444

Intercept: 0.044

Slope: 0.9286

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.148	0.11		02/24/2022	10:00
2	CAL2	1	1	1.160	1.20	-20.2	02/24/2022	10:04
3	CAL3	5	1	4.380	4.67	6.6	02/24/2022	10:07
4	CAL4	10	1	9.450	10.13	-1.3	02/24/2022	10:10
5	CAL5	20	1	18.400	19.77	1.2	02/24/2022	10:13
6	CAL6	40	1	37.300	40.12	-0.3	02/24/2022	10:17



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB118814

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.320	9.989	02/24/2022	10:20
2	ICB		1	0.136	0.099	02/24/2022	10:23
3	CCV1	10	1	9.340	10.011	02/24/2022	10:26
4	CCB1		1	0.152	0.116	02/24/2022	10:30
5	LB118814BL		1	0.156	0.121	02/24/2022	10:33
6	N1205-01		1	6.570	7.028	02/24/2022	10:36
7	N1205-01DUP		1	6.650	7.114	02/24/2022	10:40
8	N1645-07		1	0.521	0.514	02/24/2022	10:43
9	N1645-08		1	1.150	1.191	02/24/2022	10:47
10	CCV2	10	1	9.120	9.774	02/24/2022	10:50
11	CCB2		1	0.146	0.110	02/24/2022	10:54