



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118816

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH MERER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/24/2022	12:25
2	CAL2	1	Water	NA	NA	20.4	7.00	02/24/2022	12:26
3	CAL3	1	Water	NA	NA	20.4	10.03	02/24/2022	12:27
4	ICV	1	Water	NA	NA	20.7	6.99	02/24/2022	12:30
5	CCV1	1	Water	NA	NA	20.6	2.01	02/24/2022	12:31
6	N1618-01	1	Solid	20.03	20	20.5	8.17	02/24/2022	12:33
7	N1618-02	1	Solid	20.04	20	20.3	7.82	02/24/2022	12:34
8	N1618-03	1	Solid	20.04	20	20.2	7.97	02/24/2022	12:35
9	N1618-04	1	Solid	20.04	20	20.2	7.97	02/24/2022	12:35
10	N1618-05	1	Solid	20.04	20	20.2	7.97	02/24/2022	12:35
11	N1618-06	1	Solid	20.05	20	20.5	6.99	02/24/2022	12:40
12	N1618-07	1	Solid	20.03	20	20.4	6.34	02/24/2022	12:42
13	N1618-08	1	Solid	20.04	20	20.6	6.17	02/24/2022	12:45
14	N1618-09	1	Solid	20.02	20	20.1	6.17	02/24/2022	12:46
15	N1618-10	1	Solid	20.03	20	20.3	3.24	02/24/2022	12:49
16	CCV2	1	Water	NA	NA	20.5	12.03	02/24/2022	12:50

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1618 WorkList ID : 157119 Department : Wet-Chemistry Date : 02-24-2022 12:06:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/15/2022	Solid	N1618-01	pH	Cool 4 deg C	USEP01	A62	YAY22	02/15/2022	9045D
02/15/2022	Solid	N1618-02	pH	Cool 4 deg C	USEP01	A62	YAY23	02/15/2022	9045D
02/15/2022	Solid	N1618-03	pH	Cool 4 deg C	USEP01	A62	YAY24	02/15/2022	9045D
02/15/2022	Solid	N1618-04	pH	Cool 4 deg C	USEP01	A62	YAY24D	02/15/2022	9045D
02/15/2022	Solid	N1618-05	pH	Cool 4 deg C	USEP01	A62	YAY24S	02/15/2022	9045D
02/17/2022	Solid	N1618-06	pH	Cool 4 deg C	USEP01	A62	YAY25	02/17/2022	9045D
02/16/2022	Solid	N1618-07	pH	Cool 4 deg C	USEP01	A62	YAY26	02/16/2022	9045D
02/16/2022	Solid	N1618-08	pH	Cool 4 deg C	USEP01	A62	YAY27	02/16/2022	9045D
02/15/2022	Solid	N1618-09	pH	Cool 4 deg C	USEP01	A62	YAY32	02/15/2022	9045D
02/15/2022	Solid	N1618-10	pH	Cool 4 deg C	USEP01	A62	YAY33	02/15/2022	9045D

Date/Time 02/24/2022 12:11
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 02/24/2022 14:30
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)