



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118819

Slope : 99.2

pH Meter ID : WC PH MERER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.02	02/24/2022	13:15
2	CAL2	1	Water	NA	NA	20.4	7.00	02/24/2022	13:16
3	CAL3	1	Water	NA	NA	20.4	10.03	02/24/2022	13:18
4	ICV	1	Water	NA	NA	20.4	6.99	02/24/2022	13:20
5	CCV1	1	Water	NA	NA	20.6	2.01	02/24/2022	13:22
6	N1659-03	1	Water	NA	NA	20.6	6.92	02/24/2022	13:25
7	N1659-03DUP	1	Water	NA	NA	20.7	6.93	02/24/2022	13:26
8	CCV2	1	Water	NA	NA	20.5	12.03	02/24/2022	13:29

WORKLIST(Hardcopy Internal Chain)

118819

WorkList Name : PH N1659 W

WorkList ID : 157128

Department : Wet-Chemistry

Date : 02-24-2022 13:08:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/24/2022	Water	N1659-03	pH	F	Cool 4 deg C	TULL02 P11	TP-3-MH	02/24/2022	9040C

Date/Time

02/24/22 13:10

Raw Sample Received by:

JP (WC)

Raw Sample Relinquished by:

JP (WC)

Date/Time

02/24/22 13:20

Raw Sample Received by:

JP (WC)

Raw Sample Relinquished by:

JP (WC)