



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/25/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 02/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:26
Out Date: 02/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118821

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1586-11	PH-9-2.5-3	14	1.14	8.58	9.72	8.11	81.2	
N1586-21	PH-12-2.5-3	15	1.13	8.55	9.68	8.93	91.2	
N1586-21D	PH-12-2.5-3DUP	16	1.16	8.54	9.7	8.95	91.2	
N1645-01	LOD-MDL-SOIL-01-QT1-2022	1	1.00	1.00	2.00	2.00	100.0	
N1645-02	LOQ-SOIL-02-QT1-2022	2	1.00	1.00	2.00	2.00	100.0	
N1645-03	MDL-SOIL-03-QT1-2022	3	1.00	1.00	2.00	2.00	100.0	
N1645-04	LOD-MDL-SOIL-04-QT1-2022	4	1.00	1.00	2.00	2.00	100.0	
N1645-05	LOQ-SOIL-05-QT1-2022	5	1.00	1.00	2.00	2.00	100.0	
N1645-06	MDL-SOIL-06-QT1-2022	6	1.00	1.00	2.00	2.00	100.0	
N1645-16	LLOQ-SOIL-QT1-2022	7	1.00	1.00	2.00	2.00	100.0	
N1657-01	GROUP-1	38	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N1657-02	GROUP-2	39	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N1657-03	GROUP-3	40	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N1657-04	GROUP-4	41	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N1657-05	GROUP-5	42	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N1657-06	GROUP-6	43	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N1657-07	GROUP-7	44	1.00	1.00	2.00	2.00	100.0	
N1658-01	1-3-DRUMS	45	1.00	1.00	2.00	2.00	100.0	OILY DEBRIS
N1658-02	1-4-TOTES	46	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N1659-01	PIT-11-SECTION-A	8	1.14	8.78	9.92	9.16	91.3	
N1659-02	PIT-11-SECTOIN-B	9	1.12	8.75	9.87	9.06	90.7	
N1662-01	POST-EX-01-0224	10	1.15	8.81	9.96	8.73	86.0	
N1662-02	POST-EX-04-0224	11	1.13	8.47	9.6	8.44	86.3	
N1662-03	POST-EX-01B	12	1.14	8.74	9.88	8.87	88.4	
N1662-04	POST-EX-04B	13	1.16	8.62	9.78	8.35	83.4	
N1663-01	CL-02-022422	47	1.18	8.53	9.71	8.61	87.1	
N1663-02	CL-02-022422-EPH-2	48	1.11	8.72	9.83	8.84	88.6	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/25/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 02/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:26
Out Date: 02/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118821

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1664-01	TP-1	17	1.14	8.59	9.73	8.8	89.2	
N1664-03	TP-1-EPH2	18	1.14	8.52	9.66	8.74	89.2	
N1664-04	TP-2	19	1.16	8.73	9.89	9.16	91.6	
N1664-06	TP-2-EPH2	20	1.15	8.80	9.95	9.33	93.0	
N1664-07	TP-3	21	1.13	8.57	9.7	8.77	89.1	
N1664-09	TP-3-EPH3	22	1.1	8.59	9.69	8.98	91.7	
N1664-10	TP-4	23	1.11	8.65	9.76	8.72	88.0	
N1664-12	TP-4-EPH2	24	1.11	8.64	9.75	8.67	87.5	
N1664-13	TP-5	25	1.13	8.80	9.93	8.52	84.0	
N1664-15	TP-5-EPH2	26	1.11	8.61	9.72	8.6	87.0	
N1664-16	TP-6	27	1.12	8.63	9.75	9.06	92.0	
N1664-18	TP-6-EPH2	28	1.11	8.75	9.86	9.42	95.0	
N1664-19	TP-7	29	1.12	8.57	9.69	8.33	84.1	
N1664-21	TP-7-EPH2	30	1.13	8.67	9.8	8.89	89.5	
N1664-22	TP-8	31	1.1	8.58	9.68	9.14	93.7	
N1664-24	TP-8-EPH2	32	1.11	8.71	9.82	9.19	92.8	
N1664-25	TP-9	33	1.12	8.72	9.84	8.42	83.7	
N1664-27	TP-9-EPH2	34	1.1	8.88	9.98	9.44	93.9	
N1664-28	TP-10	35	1.12	8.63	9.75	8.81	89.1	
N1664-31	TP-11	36	1.1	8.42	9.52	8.73	90.6	
N1664-33	TP-11-EPH2	37	1.11	8.46	9.57	8.8	90.9	
N1669-01	BRAKER-PEDESTALS	49	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N1669-02	T2-PAD	50	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N1669-03	CAP-BANK-PEDESTALS	51	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N1670-01	OR-03-022422	52	1.18	8.61	9.79	8.99	90.7	
N1670-02	OR-03-022422-EPH-2	53	1.11	8.71	9.82	8.97	90.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$