



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB118822

BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP95453
calibration std. phosphate 0.5 ppm	WP95454
calibration std. phosphate 0.3 ppm	WP95455
calibration std. phosphate 0.1 ppm	WP95456
calibration std. phosphate 0.05 ppm	WP95457
calibration std. 0 ppm	WP95458
phosphate CCV std.	WP95469
Combined reagent	WP95460
Phenolphthalein indicator	WP93767
Sodium hydroxide, 1N	WP95405
Phosphosphate LOQ std, 0.05PPM	WP95462
Phosphate ICV-LCS Std	WP95464
Phosphate LOD-MDL Std 0.025ppm	WP95463

Intercept: -0.0019

Slope: 0.6545

Regression: 0.999955

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.003		02/24/2022	12:45
2	CAL2	0.05	1	50	50	0.031	0.05	0	02/24/2022	12:45
3	CAL3	0.10	1	50	50	0.065	0.102	-2	02/24/2022	12:46
4	CAL4	0.30	1	50	50	0.190	0.293	2.3	02/24/2022	12:47
5	CAL5	0.50	1	50	50	0.325	0.499	0.2	02/24/2022	12:47
6	CAL6	1.00	1	50	50	0.654	1.002	-0.2	02/24/2022	12:48



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB118822

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.501	02/24/2022	12:49
2	ICB		1	50	50	0.002	0.006	02/24/2022	12:49
3	CCV1	0.50	1	50	50	0.322	0.495	02/24/2022	13:05
4	CCB1		1	50	50	0.003	0.007	02/24/2022	13:05
5	PB142904BL		1	1	50	0.004	0.009	02/24/2022	13:06
6	PB142904BS	0.50	1	1	50	0.328	0.504	02/24/2022	13:06
7	N1645-01		1	1	50	0.015	0.026	02/24/2022	13:07
8	N1645-02		1	1	50	0.032	0.052	02/24/2022	13:08
9	CCV2	0.50	1	1	50	0.325	0.499	02/24/2022	13:08
10	CCB2		1	1	50	0.003	0.007	02/24/2022	13:09