


LB118423

7470A

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.5999e-004

C= 1.0615e-002

Rho= 0.9982427

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/r
0.0	0.000	0.017	0.017	24	0.000	24					—
0.2	0.200	0.237	0.037	872	0.0 %	872					19
2.5	2.500	2.243	-0.257	8588	0.0 %	8588					-10
5.0	5.000	5.175	0.175	19864	0.0 %	19864					4
7.5	7.500	7.821	0.321	30039	0.0 %	30039					4
10.0	10.000	9.707	-0.293	37294	0.0 %	37294					-9

LB118823 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Feb 2022 15:21:50

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	7470A	PPB	24 Feb 2022 15:22:01	S	Std
0.2 - 1	51.2	872	-	0.2000	7470A	PPB	24 Feb 2022 15:24:19	S	Std
2.5 - 1	52.5	8588	-	2.5000	7470A	PPB	24 Feb 2022 15:26:38	S	Std
5.0 - 1	55	19864	-	5.0000	7470A	PPB	24 Feb 2022 15:28:55	S	Std
7.5 - 1	57.5	30039	-	7.5000	7470A	PPB	24 Feb 2022 15:31:12	S	Std
10.0 - 1	58	37294	-	10.0000	7470A	PPB	24 Feb 2022 15:38:49	S	Std
ICV06 - 1	ICV06	14396	3.7535	-	7470A	PPB	24 Feb 2022 15:42:05	U	SMPL
ICB06 - 1	ICB06	-2	0.0101	-	7470A	PPB	24 Feb 2022 15:44:20	U	SMPL
CCV46 - 1	CCV46	20058	5.2255	-	7470A	PPB	24 Feb 2022 15:46:36	U	SMPL
CCB46 - 1	CCB46	-22	0.0049	-	7470A	PPB	24 Feb 2022 15:48:51	U	SMPL
CRA - 1	CRA	831	0.2267	-	7470A	PPB	24 Feb 2022 15:51:06	U	SMPL
HighStd - 1	HighStd	37634	9.7952	-	7470A	PPB	24 Feb 2022 15:53:21	U	SMPL
ChkStd - 1	ChkStd	27516	7.1646	-	7470A	PPB	24 Feb 2022 15:58:16	U	SMPL
PB142905BL - 1	PBW	-31	0.0026	-	7470A	PPB	24 Feb 2022 16:00:31	U	SMPL
PB142905BS - 1	LCSW	14849	3.8712	-	7470A	PPB	24 Feb 2022 16:02:47	U	SMPL
N1639-02 - 1	WP-SOIL-1-FINAL	17	0.0150	-	7470A	PPB	24 Feb 2022 16:05:04	U	SMPL
N1639-02DUP - 1	WP-SOIL-1-FINALDUP	47	0.0228	-	7470A	PPB	24 Feb 2022 16:07:20	U	SMPL
N1639-02MS - 1	WP-SOIL-1-FINALMS	16773	4.3715	-	7470A	PPB	24 Feb 2022 16:09:37	U	SMPL
N1639-02MSD - 1	WP-SOIL-1-FINALMSD	16240	4.2329	-	7470A	PPB	24 Feb 2022 16:11:54	U	SMPL
N1650-02 - 1	SW-SURFICIAL-SOIL-(SW-SS)	-5	0.0093	-	7470A	PPB	24 Feb 2022 16:14:10	U	SMPL
CCV47 - 1	CCV47	17577	4.5805	-	7470A	PPB	24 Feb 2022 16:19:02	U	SMPL
CCB47 - 1	CCB47	-40	0.0002	-	7470A	PPB	24 Feb 2022 16:21:17	U	SMPL
N1650-04 - 1	NW-SURFICIAL-SOIL-(NW-SS)	230	0.0704	-	7470A	PPB	24 Feb 2022 16:23:32	U	SMPL
PB142917BL - 1	PBW	13	0.0140	-	7470A	PPB	24 Feb 2022 16:25:48	U	SMPL
PB142917BS - 1	LCSW	16326	4.2552	-	7470A	PPB	24 Feb 2022 16:28:03	U	SMPL
N1654-01 - 1	20220091-BERKELEY-3-ELUTIRATE	55	0.0249	-	7470A	PPB	24 Feb 2022 16:30:19	U	SMPL
N1654-01DUP - 1	20220091-BERKELEY-3-ELUTIRATEDUP	90	0.0340	-	7470A	PPB	24 Feb 2022 16:32:35	U	SMPL
N1654-01MS - 1	20220091-BERKELEY-3-ELUTIRATEMS	14541	3.7912	-	7470A	PPB	24 Feb 2022 16:34:52	U	SMPL
N1654-01MSD - 1	20220091-BERKELEY-3-ELUTIRATEMSD	15161	3.9524	-	7470A	PPB	24 Feb 2022 16:37:08	U	SMPL
PB142887TB - 1	PB142887TB	-44	-0.0008	-	7470A	PPB	24 Feb 2022 16:39:26	U	SMPL
CCV48 - 1	CCV48	20001	5.2107	-	7470A	PPB	24 Feb 2022 16:41:44	U	SMPL
CCB48 - 1	CCB48	-41	-0.0000	-	7470A	PPB	24 Feb 2022 16:43:59	U	SMPL
N1639-02LX5 - 1		16	0.0148	-	7470A	PPB	24 Feb 2022 16:46:15	U	SMPL
N1639-02A - 1		15650	4.0795	-	7470A	PPB	24 Feb 2022 16:48:31	U	SMPL
N1654-01LX5 - 1		-19	0.0057	-	7470A	PPB	24 Feb 2022 16:50:46	U	SMPL
N1654-01A - 1		15685	4.0886	-	7470A	PPB	24 Feb 2022 16:53:02	U	SMPL
CCV49 - 1	CCV49	20384	5.3103	-	7470A	PPB	24 Feb 2022 16:55:18	U	SMPL
CCB49 - 1	CCB49	-39	0.0005	-	7470A	PPB	24 Feb 2022 16:57:34	U	SMPL