



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 02/25/2022OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 02/24/2022 17:40 DATE & TIME OUT1: 02/25/2022 08:00OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 02/25/2022 08:30 DATE & TIME OUT2: 02/25/2022 10:00Run Number: LB118831BalanceID: WC SC-6OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB118831BL	LB118831BL	75.6391	75.6391	100	75.6393	75.6393	75.6393	0.0002	2
2	N1668-01	PURGED-WATER	68.8127	68.8127	100	68.9740	68.9740	68.9740	0.1613	1613
3	N1668-01DUP	PURGED-WATERDUP	84.7433	84.7433	100	84.9114	84.9114	84.9114	0.1681	1681

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

157118231

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSWN1669 WorkList ID : 157145 Department : Wet-Chemistry Date : 02-24-2022 17:28:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
03/03/2022	Water	N1668-01	TS	Cool 4 deg C	PSEG03	M11	PURGED-WATER	02/24/2022	SM2540 B

Date/Time 02/24/22 16:10
Raw Sample Received by: JD COG
Raw Sample Relinquished by: ad sm

Date/Time 02/24/22 18:00
Raw Sample Received by: CD SM
Raw Sample Relinquished by: JD COG