



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB118841

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP95487
Turbidity Calibration std, 0NTU	WP95486
Turbidity Calibration std, 20NTU	WP95481
Turbidity Calibration std, 40NTU	WP95480
Turbidity Calibration std, 1NTU	WP95484
Turbidity Calibration std, 5NTU	WP95483
Turbidity Calibration - CCV std, 10 NTU	WP95482

Intercept: 0.0359

Slope: 0.9315

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.142	0.11		02/25/2022	09:00
2	CAL2	1	1	1.154	1.20	-20	02/25/2022	09:03
3	CAL3	5	1	4.450	4.74	5.2	02/25/2022	09:06
4	CAL4	10	1	9.560	10.22	-2.2	02/25/2022	09:10
5	CAL5	20	1	18.200	19.50	2.5	02/25/2022	09:13
6	CAL6	40	1	37.500	40.22	-0.5	02/25/2022	09:16



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB118841

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.440	10.096	02/25/2022	09:19
2	ICB		1	0.132	0.103	02/25/2022	09:21
3	CCV1	10	1	9.520	10.182	02/25/2022	09:24
4	CCB1		1	0.142	0.114	02/25/2022	09:26
5	LB118841BL		1	0.158	0.131	02/25/2022	09:28
6	N1645-09		1	0.533	0.534	02/25/2022	09:30
7	CCV2	10	1	9.220	9.859	02/25/2022	09:46
8	CCB2		1	0.152	0.125	02/25/2022	09:50