



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB118842

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP95487
Turbidity Calibration std, 0NTU	WP95486
Turbidity Calibration std, 20NTU	WP95481
Turbidity Calibration std, 40NTU	WP95480
Turbidity Calibration std, 1NTU	WP95484
Turbidity Calibration std, 5NTU	WP95483
Turbidity Calibration - CCV std, 10 NTU	WP95482

Intercept: 0.0359

Slope: 0.9315

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.142	0.11		02/25/2022	09:00
2	CAL2	1	1	1.154	1.20	-20	02/25/2022	09:03
3	CAL3	5	1	4.450	4.74	5.2	02/25/2022	09:06
4	CAL4	10	1	9.560	10.22	-2.2	02/25/2022	09:10
5	CAL5	20	1	18.200	19.50	2.5	02/25/2022	09:13
6	CAL6	40	1	37.500	40.22	-0.5	02/25/2022	09:16



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB118842

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.440	10.096	02/25/2022	09:19
2	ICB		1	0.132	0.103	02/25/2022	09:21
3	CCV1	10	1	9.520	10.182	02/25/2022	09:24
4	CCB1		1	0.142	0.114	02/25/2022	09:26
5	LB118842BL		1	0.158	0.131	02/25/2022	09:32
6	N1645-12		1	0.533	0.534	02/25/2022	09:35
7	N1671-01		1	0.240	0.219	02/25/2022	09:28
8	N1671-01DUP		1	0.242	0.221	02/25/2022	09:30
9	N1671-03		1	17.300	18.534	02/25/2022	09:37
10	N1671-05		1	0.511	0.510	02/25/2022	09:40
11	N1671-07		1	15.200	16.279	02/25/2022	09:43
12	CCV2	10	1	9.220	9.859	02/25/2022	09:46
13	CCB2		1	0.152	0.125	02/25/2022	09:50

L13118841
L13118842

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TURB-022522 WorkList ID : 157161 Department : Wet-Chemistry Date : 02-25-2022 09:11:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage Location	Customer Sample		
02/25/2022	Water	N1645-09	Turbidity	Cool 4 deg C	CHEM02	QA Of	MDL-WATER-03-QT1-2022	02/23/2022	180.1
02/25/2022	Water	N1645-12	Turbidity	Cool 4 deg C	CHEM02	QA Of	MDL-WATER-06-QT1-2022	02/23/2022	SM2130 B
02/25/2022	Water	N1671-01	Turbidity	Cool 4 deg C	LOCK01	N31	MW-1	02/23/2022	SM2130 B
02/25/2022	Water	N1671-03	Turbidity	Cool 4 deg C	LOCK01	N31	MW-2	02/23/2022	SM2130 B
02/25/2022	Water	N1671-05	Turbidity	Cool 4 deg C	LOCK01	N31	MW-3	02/23/2022	SM2130 B
02/25/2022	Water	N1671-07	Turbidity	Cool 4 deg C	LOCK01	N31	MW-4	02/23/2022	SM2130 B

Date/Time 02/25/22 8:45
Raw Sample Received by: 12 (20)
Raw Sample Relinquished by: JH (2007)

Date/Time 02/25/22 10:10
Raw Sample Received by: JH (2007)
Raw Sample Relinquished by: 12 (20)