



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118846

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	02/25/2022	10:45
2	CAL2	1	Water	NA	NA	20.6	7.00	02/25/2022	10:46
3	CAL3	1	Water	NA	NA	20.6	10.03	02/25/2022	10:47
4	ICV	1	Water	NA	NA	20.6	6.98	02/25/2022	10:48
5	CCV1	1	Water	NA	NA	20.6	2.01	02/25/2022	10:50
6	N1668-01	1	Water	NA	NA	20.2	6.58	02/25/2022	11:00
7	N1671-01	1	Water	NA	NA	20.2	5.90	02/25/2022	11:05
8	N1671-03	1	Water	NA	NA	20.3	6.80	02/25/2022	11:06
9	N1671-05	1	Water	NA	NA	20.4	6.68	02/25/2022	11:10
10	N1671-07	1	Water	NA	NA	20.4	6.97	02/25/2022	11:15
11	N1671-07DUP	1	Water	NA	NA	20.5	6.98	02/25/2022	11:16
12	CCV2	1	Water	NA	NA	20.7	12.03	02/25/2022	11:18

W2118846

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1671 WorkList ID : 157174 Department : Wet-Chemistry Date : 02-25-2022 10:38:07

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample		
				Preservative	Customer	Storage Location	Location	Collect Date	Method
02/24/2022	Water	N1668-01	pH	Cool 4 deg C	PSEG03	M11	PURGED-WATER	02/24/2022	9040C
02/23/2022	Water	N1671-01	pH	Cool 4 deg C	LOCK01	N31	MW-1	02/23/2022	9040C
02/23/2022	Water	N1671-03	pH	Cool 4 deg C	LOCK01	N31	MW-2	02/23/2022	9040C
02/23/2022	Water	N1671-05	pH	Cool 4 deg C	LOCK01	N31	MW-3	02/23/2022	9040C
02/23/2022	Water	N1671-07	pH	Cool 4 deg C	LOCK01	N31	MW-4	02/23/2022	9040C

Date/Time 02/25/22 10:40
Raw Sample Received by: JDLWC
Raw Sample Relinquished by: KS(WC)

Date/Time 02/25/22 12:00
Raw Sample Received by: KS(WC)
Raw Sample Relinquished by: JDLWC