

LB118849


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45487



A= 0.0000e+000

B= 2.1515e-004

C= 6.5994e-002

Rho= 0.9984841

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0.01
0.0	0.000	0.109	0.109	201	0.000	201					1
0.2	0.200	0.235	0.035	785	0.0 %	785					18
2.5	2.500	2.133	-0.367	9609	0.0 %	9609					-15
5.0	5.000	5.255	0.255	24119	0.0 %	24119					5
7.5	7.500	7.614	0.114	35084	0.0 %	35084					2
10.0	10.000	9.853	-0.147	45487	0.0 %	45487					-1

LB118849 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 25 Feb 2022 11:45:30					
Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	201	-	0.0000	7471B	PPB	25 Feb 2022 11:46:40	S	Std
0.2 - 1	50-2	785	-	0.2000	7471B	PPB	25 Feb 2022 11:48:57	S	Std
2.5 - 1	52.5	9609	-	2.5000	7471B	PPB	25 Feb 2022 11:51:14	S	Std
5.0 - 1	55	24119	-	5.0000	7471B	PPB	25 Feb 2022 11:53:32	S	Std
7.5 - 1	57.5	35084	-	7.5000	7471B	PPB	25 Feb 2022 11:55:49	S	Std
10.0 - 1	51	45487	-	10.0000	7471B	PPB	25 Feb 2022 12:03:25	S	Std
ICV07 - 1	ICV07	18877	4.1275	-	7471B	PPB	25 Feb 2022 12:08:40	U	SMPL
ICB07 - 1	ICB07	-19	0.0619	-	7471B	PPB	25 Feb 2022 12:10:55	U	SMPL
CCV46 - 1	CCV46	23778	5.1819	-	7471B	PPB	25 Feb 2022 12:13:12	U	SMPL
CCB46 - 1	CCB46	-42	0.0570	-	7471B	PPB	25 Feb 2022 12:15:27	U	SMPL
CRA - 1	CRA	789	0.2358	-	7471B	PPB	25 Feb 2022 12:20:19	U	SMPL
HighStd - 1	HighStd	49040	10.6171	-	7471B	PPB	25 Feb 2022 12:22:34	U	SMPL
ChkStd - 1	ChkStd	33472	7.2676	-	7471B	PPB	25 Feb 2022 12:24:49	U	SMPL
PB142940BL - 1	PBS	21	0.0705	-	7471B	PPB	25 Feb 2022 12:27:07	U	SMPL
PB142940BS - 1	LCSS	18878	4.1277	-	7471B	PPB	25 Feb 2022 12:29:24	U	SMPL
N1650-01 - 1	SW-SURFICIAL-SOIL-(SW-SS)	68959	14.9028	-	7471B	PPB	25 Feb 2022 12:31:41	U	SMPL
N1650-03 - 1	NW-SURFICIAL-SOIL-(NW-SS)	225236	48.5264	-	7471B	PPB	25 Feb 2022 12:33:57	U	SMPL
N1658-01 - 1	1-3-DRUMS	-276	0.0066	-	7471B	PPB	25 Feb 2022 12:36:16	U	SMPL
N1662-02 - 1	POST-EX-04-0224	59693	12.9092	-	7471B	PPB	25 Feb 2022 12:38:37	U	SMPL
N1662-04 - 1	POST-EX-04B	91081	19.6624	-	7471B	PPB	25 Feb 2022 12:40:52	U	SMPL
CCV47 - 1	CCV47	23844	5.1961	-	7471B	PPB	25 Feb 2022 12:45:45	U	SMPL
CCB47 - 1	CCB47	157	0.0998	-	7471B	PPB	25 Feb 2022 12:48:00	U	SMPL
N1663-01 - 1	CL-02-022422	5521	1.2539	-	7471B	PPB	25 Feb 2022 12:50:16	U	SMPL
N1669-01 - 1	BRAKER-PEDESTALS	370	0.1456	-	7471B	PPB	25 Feb 2022 12:52:31	U	SMPL
N1669-02 - 1	T2-PAD	351	0.1415	-	7471B	PPB	25 Feb 2022 12:54:47	U	SMPL
N1669-03 - 1	CAP-BANK-PEDESTALS	311	0.1329	-	7471B	PPB	25 Feb 2022 12:57:03	U	SMPL
N1670-01 - 1	OR-03-022422	3540	0.8276	-	7471B	PPB	25 Feb 2022 12:59:19	U	SMPL
N1670-01MSD - 1	OR-03-022422MSD	23204	5.0584	-	7471B	PPB	25 Feb 2022 13:06:10	U	SMPL
PB142941BL - 1	PBS	60	0.0789	-	7471B	PPB	25 Feb 2022 13:08:27	U	SMPL
PB142941BS - 1	LCSS	19903	4.3482	-	7471B	PPB	25 Feb 2022 13:10:43	U	SMPL
CCV48 - 1	CCV48	22106	4.8222	-	7471B	PPB	25 Feb 2022 13:12:59	U	SMPL
CCB48 - 1	CCB48	-6	0.0647	-	7471B	PPB	25 Feb 2022 13:15:15	U	SMPL
N1664-01 - 1	TP-1	1545	0.3984	-	7471B	PPB	25 Feb 2022 13:22:35	U	SMPL
N1664-01DUP - 1	TP-1DUP	1547	0.3988	-	7471B	PPB	25 Feb 2022 13:24:51	U	SMPL
N1664-01MS - 1	TP-1MS	22779	4.9670	-	7471B	PPB	25 Feb 2022 13:27:06	U	SMPL
N1664-01MSD - 1	TP-1MSD	21526	4.6974	-	7471B	PPB	25 Feb 2022 13:31:58	U	SMPL
N1664-04 - 1	TP-2	1614	0.4133	-	7471B	PPB	25 Feb 2022 13:34:14	U	SMPL
N1664-07 - 1	TP-3	36681	7.9581	-	7471B	PPB	25 Feb 2022 13:36:31	U	SMPL
N1664-10 - 1	TP-4	20601	4.4984	-	7471B	PPB	25 Feb 2022 13:38:48	U	SMPL
N1664-13 - 1	TP-5	66163	14.3012	-	7471B	PPB	25 Feb 2022 13:41:06	U	SMPL
N1664-16 - 1	TP-6	553	0.1850	-	7471B	PPB	25 Feb 2022 13:43:23	U	SMPL
N1664-19 - 1	TP-7	39607	8.5876	-	7471B	PPB	25 Feb 2022 13:45:40	U	SMPL
CCV49 - 1	CCV49	24419	5.3198	-	7471B	PPB	25 Feb 2022 13:47:56	U	SMPL
CCB49 - 1	CCB49	-7	0.0645	-	7471B	PPB	25 Feb 2022 13:50:12	U	SMPL
N1664-22 - 1	TP-8	1374	0.3616	-	7471B	PPB	25 Feb 2022 13:52:28	U	SMPL
N1664-25 - 1	TP-9	56460	12.2136	-	7471B	PPB	25 Feb 2022 13:54:44	U	SMPL
N1664-28 - 1	TP-10	3401	0.7977	-	7471B	PPB	25 Feb 2022 13:57:00	U	SMPL
N1664-31 - 1	TP-11	5961	1.3485	-	7471B	PPB	25 Feb 2022 13:59:17	U	SMPL
N1670-01DUP - 1	OR-03-022422DUP	3530	0.8255	-	7471B	PPB	25 Feb 2022 14:01:33	U	SMPL
N1670-01MS - 1	OR-03-022422MS	23515	5.1253	-	7471B	PPB	25 Feb 2022 14:03:49	U	SMPL
N1670-01LX5 - 1		610	0.1972	-	7471B	PPB	25 Feb 2022 14:06:06	U	SMPL
N1670-01A - 1		23505	5.1232	-	7471B	PPB	25 Feb 2022 14:08:23	U	SMPL
N1664-01LX5 - 1		306	0.1318	-	7471B	PPB	25 Feb 2022 14:10:40	U	SMPL
N1664-01A - 1		21260	4.6402	-	7471B	PPB	25 Feb 2022 14:12:56	U	SMPL
CCV50 - 1	CCV50	24808	5.4035	-	7471B	PPB	25 Feb 2022 14:15:12	U	SMPL
CCB50 - 1	CCB50	-14	0.0630	-	7471B	PPB	25 Feb 2022 14:17:28	U	SMPL
N1650-01DLX5 - 1		15297	3.3572	-	7471B	PPB	25 Feb 2022 14:19:43	U	SMPL
N1662-02DLX2 - 1		32085	6.9692	-	7471B	PPB	25 Feb 2022 14:24:15	U	SMPL
N1662-04DLX5 - 1		20440	4.4637	-	7471B	PPB	25 Feb 2022 14:26:31	U	SMPL
N1664-13DLX5 - 1		14270	3.1362	-	7471B	PPB	25 Feb 2022 14:28:47	U	SMPL
N1664-25DLX2 - 1		28909	6.2859	-	7471B	PPB	25 Feb 2022 14:31:03	U	SMPL
N1650-03DLX10 - 1		26929	5.8599	-	7471B	PPB	25 Feb 2022 14:33:40	U	SMPL
CCV51 - 1	CCV51	23539	5.1305	-	7471B	PPB	25 Feb 2022 14:36:23	U	SMPL
CCB51 - 1	CCB51	27	0.0718	-	7471B	PPB	25 Feb 2022 14:38:40	U	SMPL