



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118852

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/25/2022	12:50
2	CAL2	1	Water	NA	NA	20.3	7.00	02/25/2022	12:51
3	CAL3	1	Water	NA	NA	20.3	10.03	02/25/2022	12:52
4	ICV	1	Water	NA	NA	20.4	6.99	02/25/2022	12:55
5	CCV1	1	Water	NA	NA	20.3	2.02	02/25/2022	12:56
6	N1680-01	1	Water	NA	NA	20.2	4.44	02/25/2022	13:00
7	N1680-01DUP	1	Water	NA	NA	20.3	4.44	02/25/2022	13:01
8	N1680-02	1	Water	NA	NA	20.2	4.44	02/25/2022	13:02
9	N1680-03	1	Water	NA	NA	20.1	4.42	02/25/2022	13:05
10	CCV2	1	Water	NA	NA	20.5	12.03	02/25/2022	13:06

WB118852

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1680 WorkList ID : 157182 Department : Wet-Chemistry Date : 02-25-2022 12:42:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/23/2022	Water	N1680-01	pH	Cool 4 deg C	STER13	N11	CELL-A-POOL-WATER	02/23/2022	9040C
02/23/2022	Water	N1680-02	pH	Cool 4 deg C	STER13	N11	CELL-B-POOL-WATER	02/23/2022	9040C
02/23/2022	Water	N1680-03	pH	Cool 4 deg C	STER13	N11	CELL-C-POOL-WATER	02/23/2022	9040C

Date/Time 02/25/22 12:45
 Raw Sample Received by: JP (W/C)
 Raw Sample Relinquished by: JP (W/C)

Date/Time 02/25/22
 Raw Sample Received by: JP (W/C)
 Raw Sample Relinquished by: JP (W/C)