



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/1/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 15:30
In Date: 02/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 02/26/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118859

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1682-01	MDBRJ5	1	1.18	8.49	9.67	8.00	80.3	
N1682-02	MDBRJ5D	2	1.18	8.49	9.67	8.00	80.3	
N1682-03	MDBRJ5S	3	1.18	8.49	9.67	8.00	80.3	
N1682-04	MDBRJ6	4	1.19	8.69	9.88	8.02	78.6	
N1682-05	MDBRJ7	5	1.18	8.48	9.66	7.79	77.9	
N1682-06	MDBRJ8	6	1.15	8.82	9.97	8.51	83.4	
N1682-07	MDBRJ9	7	1.16	8.52	9.68	7.89	79.0	
N1682-08	MDBRK0	8	1.17	8.74	9.91	8.26	81.1	
N1682-09	MDBRK1	9	1.19	8.40	9.59	8.13	82.6	
N1682-10	MDBRK2	10	1.18	8.61	9.79	7.92	78.3	
N1682-11	MDBRK3	11	1.18	8.79	9.97	8.36	81.7	
N1682-14	MDBRL3	12	1.15	8.76	9.91	8.29	81.5	
N1682-15	MDBRL4	13	1.15	8.81	9.96	8.24	80.5	
N1682-16	MDBRL5	14	1.17	8.65	9.82	8.3	82.4	
N1682-17	MDBRL6	15	1.12	8.67	9.79	8.12	80.7	
N1682-18	MDBRL7	16	1.17	8.69	9.86	8.28	81.8	
N1682-19	MDBRL9	17	1.16	8.54	9.7	7.99	80.0	
N1682-20	MDBRM0	18	1.16	8.81	9.97	8.65	85.0	
N1682-21	MDBRM1	19	1.18	8.60	9.78	8.15	81.0	
N1682-22	MDBRM2	20	1.17	8.74	9.91	8.26	81.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$