



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/28/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 02/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:32
Out Date: 02/26/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118867

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1684-01	MDBRK4	1	1.14	8.41	9.55	8.59	88.6	
N1684-02	MDBRK4D	2	1.14	8.41	9.55	8.59	88.6	
N1684-03	MDBRK4S	3	1.14	8.41	9.55	8.59	88.6	
N1684-04	MDBRK5	4	1.18	8.35	9.53	8.68	89.8	
N1684-05	MDBRK6	5	1.19	8.39	9.58	9.57	99.9	
N1684-06	MDBRK7	6	1.15	8.60	9.75	8.94	90.6	
N1684-07	MDBRK8	7	1.19	8.52	9.71	8.48	85.6	
N1684-08	MDBRK9	8	1.18	8.59	9.77	8.62	86.6	
N1684-09	MDBRM4	9	1.15	8.82	9.97	8.36	81.7	
N1684-10	MDBRM5	10	1.19	8.62	9.81	7.9	77.8	
N1684-11	MDBRM7	11	1.15	8.83	9.98	8.36	81.7	
N1684-12	MDBRM8	12	1.12	8.44	9.56	7.78	78.9	
N1684-13	MDBRM9	13	1.15	8.81	9.96	8.1	78.9	
N1684-14	MDBRN0	14	1.15	8.83	9.98	8.37	81.8	
N1684-15	MDBRN1	15	1.16	8.62	9.78	8.11	80.6	
N1684-16	MDBRN2	16	1.19	8.63	9.82	7.82	76.8	
N1684-17	MDBRN3	17	1.15	8.83	9.98	8.66	85.1	
N1684-18	MDBRN4	18	1.14	8.81	9.95	8.06	78.5	
N1684-19	MDBRN5	19	1.19	8.51	9.7	7.78	77.4	
N1684-20	MDBRN6	20	1.19	8.78	9.97	8.54	83.7	
N1684-21	MDBRN7	21	1.18	8.77	9.95	8.11	79.0	
N1684-22	MDBRN8	22	1.16	8.82	9.98	8.39	82.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$