


L3118272

SFAM01.1

INSTRUMENT ID: LV1



Linear

A= 0.0000e+000
 B= 1.9617e-004 *Slope*
 C= -5.9619e-002 *y-intercept*
 Rho= 0.9999520
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.008	0.008	347	0.000	347					0
0.05	0.050					185	0				0
0.20	0.200	0.200	0.000	1324	0.0 %	1324					-1
2.50	2.500	2.483	-0.017	12963	0.0 %	12963					1
5.00	5.000	5.043	0.043	26013	0.0 %	26013					-1
7.50	7.500	7.432	-0.068	38192	0.0 %	38192					0
10.0	10.000	10.032	0.032	51446	0.0 %	51446					0

LB118872
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	347	-		0 SFAM01.1	PPB	2/26/2022 12:03	Std
	0.2 S01	1324	-		0.2 SFAM01.1	PPB	2/26/2022 12:05	Std
	2.5 S02	12963	-		2.5 SFAM01.1	PPB	2/26/2022 12:07	Std
	5 S03	26013	-		5 SFAM01.1	PPB	2/26/2022 12:10	Std
	7.5 S04	38192	-		7.5 SFAM01.1	PPB	2/26/2022 12:12	Std
	10 S05	51446	-		10 SFAM01.1	PPB	2/26/2022 12:14	Std
ICV008	ICV008	18716	3.6119	-	SFAM01.1	PPB	2/26/2022 12:17	SMPL
ICB008	ICB008	313	0.0018	-	SFAM01.1	PPB	2/26/2022 12:19	SMPL
CCV052	CCV052	24933	4.8314	-	SFAM01.1	PPB	2/26/2022 12:21	SMPL
CCB052	CCB052	294	-0.0019	-	SFAM01.1	PPB	2/26/2022 12:24	SMPL
PB142635BL	PBS635	419	0.0226	-	SFAM01.1	PPB	2/26/2022 12:26	SMPL
N1452-01	YAWS7	3137	0.5558	-	SFAM01.1	PPB	2/26/2022 12:33	SMPL
N1452-02	YAWS7D	3110	0.5505	-	SFAM01.1	PPB	2/26/2022 12:35	SMPL
N1452-03	YAWS7S	16040	3.0869	-	SFAM01.1	PPB	2/26/2022 12:38	SMPL
N1452-04	YAWS8	1015	0.1395	-	SFAM01.1	PPB	2/26/2022 12:40	SMPL
N1452-05	YAWS9	2198	0.3716	-	SFAM01.1	PPB	2/26/2022 12:42	SMPL
N1452-06	YAWT0	1847	0.3027	-	SFAM01.1	PPB	2/26/2022 12:44	SMPL
N1452-07	YAWT2	2599	0.4502	-	SFAM01.1	PPB	2/26/2022 12:47	SMPL
N1452-08	YAWT3	2105	0.3533	-	SFAM01.1	PPB	2/26/2022 12:49	SMPL
N1452-09	YAWT4	2834	0.4963	-	SFAM01.1	PPB	2/26/2022 12:51	SMPL
N1452-10	YAWT5	3716	0.6693	-	SFAM01.1	PPB	2/26/2022 12:53	SMPL
N1452-11	YAWT6	3493	0.6256	-	SFAM01.1	PPB	2/26/2022 12:56	SMPL
N1452-12	YAWT7	3338	0.5952	-	SFAM01.1	PPB	2/26/2022 12:58	SMPL
N1452-13	YAWT8	2899	0.5091	-	SFAM01.1	PPB	2/26/2022 13:00	SMPL
N1452-14	YAWT9	3519	0.6307	-	SFAM01.1	PPB	2/26/2022 13:03	SMPL
N1452-15	YAWW2	2028	0.3382	-	SFAM01.1	PPB	2/26/2022 13:05	SMPL
N1452-16	YAWW3	3453	0.6177	-	SFAM01.1	PPB	2/26/2022 13:07	SMPL
N1452-17	YAWW4	6027	1.1227	-	SFAM01.1	PPB	2/26/2022 13:09	SMPL
N1452-18	YAWW5	9714	1.846	-	SFAM01.1	PPB	2/26/2022 13:12	SMPL
N1452-19	YAWW7	7290	1.3704	-	SFAM01.1	PPB	2/26/2022 13:14	SMPL
CCV053	CCV053	24081	4.6643	-	SFAM01.1	PPB	2/26/2022 13:16	SMPL
CCB053	CCB053	270	-0.0067	-	SFAM01.1	PPB	2/26/2022 13:18	SMPL
N1452-20	YAWW8	5354	0.9907	-	SFAM01.1	PPB	2/26/2022 13:21	SMPL
N1452-21	YAWW9	7204	1.3536	-	SFAM01.1	PPB	2/26/2022 13:23	SMPL
PB142636BL	PBS636	176	-0.0251	-	SFAM01.1	PPB	2/26/2022 13:25	SMPL
N1453-01	YAWW6	3330	0.5936	-	SFAM01.1	PPB	2/26/2022 13:28	SMPL
N1453-02	YAWW6D	3759	0.6778	-	SFAM01.1	PPB	2/26/2022 13:30	SMPL
N1453-03	YAWW6S	15522	2.9853	-	SFAM01.1	PPB	2/26/2022 13:32	SMPL
N1453-04	YAWX0	5155	0.9516	-	SFAM01.1	PPB	2/26/2022 13:34	SMPL
N1453-05	YAWX1	5792	1.0766	-	SFAM01.1	PPB	2/26/2022 13:37	SMPL
N1453-06	YAWX2	3287	0.5852	-	SFAM01.1	PPB	2/26/2022 13:39	SMPL
N1453-07	YAWX3	16470	3.1713	-	SFAM01.1	PPB	2/26/2022 13:41	SMPL
N1453-08	YAWX4	18456	3.5609	-	SFAM01.1	PPB	2/26/2022 13:43	SMPL
N1453-09	YAWX5	331091	HIGH	-	SFAM01.1	PPB	2/26/2022 13:46	
N1453-10	YAWX6	1151	0.1662	-	SFAM01.1	PPB	2/26/2022 13:48	SMPL
N1453-11	YAWX7	726	0.0828	-	SFAM01.1	PPB	2/26/2022 13:50	SMPL

LB118872
INSTRUMENT ID : CV1

N1453-12	YAWX8	604	0.0589 -	SFAM01.1	PPB	2/26/2022 13:52	SMPL
N1453-13	YAWX9	716	0.0808 -	SFAM01.1	PPB	2/26/2022 13:54	SMPL
N1453-14	YAWY0	859	0.1089 -	SFAM01.1	PPB	2/26/2022 13:57	SMPL
N1453-15	YAWY1	1052	0.1467 -	SFAM01.1	PPB	2/26/2022 13:59	SMPL
N1453-16	YAWY2	1313	0.1979 -	SFAM01.1	PPB	2/26/2022 14:01	SMPL
N1453-17	YAWY3	143769	28.1432 -	SFAM01.1	PPB	2/26/2022 14:03	SMPL
CCV054	CCV054	24885	4.822 -	SFAM01.1	PPB	2/26/2022 14:08	SMPL
CCB054	CCB054	231	-0.0143 -	SFAM01.1	PPB	2/26/2022 14:11	SMPL
N1453-18	YAWY4	4264	0.7768 -	SFAM01.1	PPB	2/26/2022 14:13	SMPL
N1453-19	YAWY6	2409	0.4129 -	SFAM01.1	PPB	2/26/2022 14:15	SMPL
N1453-20	YAWY7	52084	10.1576 -	SFAM01.1	PPB	2/26/2022 14:18	SMPL
N1453-21	YAWY8	3350	0.5975 -	SFAM01.1	PPB	2/26/2022 14:20	SMPL
N1453-22	YAWY9	9055	1.7167 -	SFAM01.1	PPB	2/26/2022 14:22	SMPL
PB142637BL	PBS637	134	-0.0333 -	SFAM01.1	PPB	2/26/2022 14:24	SMPL
N1454-01	YAWY5	2988	0.5265 -	SFAM01.1	PPB	2/26/2022 14:27	SMPL
N1454-02	YAWY5D	2695	0.4691 -	SFAM01.1	PPB	2/26/2022 14:29	SMPL
N1454-03	YAWY5S	15535	2.9878 -	SFAM01.1	PPB	2/26/2022 14:34	SMPL
N1454-04	YAWZ0	85008	16.6162 -	SFAM01.1	PPB	2/26/2022 14:36	SMPL
N1454-05	YAWZ1	65165	12.7237 -	SFAM01.1	PPB	2/26/2022 14:38	SMPL
N1454-06	YAWZ3	4002	0.7254 -	SFAM01.1	PPB	2/26/2022 14:41	SMPL
N1454-07	YAX62	2978	0.5246 -	SFAM01.1	PPB	2/26/2022 14:43	SMPL
N1454-08	YAX63	15502	2.9814 -	SFAM01.1	PPB	2/26/2022 14:45	SMPL
N1454-09	YAX64	14219	2.7297 -	SFAM01.1	PPB	2/26/2022 14:48	SMPL
N1454-10	YAX70	4145	0.7535 -	SFAM01.1	PPB	2/26/2022 14:50	SMPL
N1454-11	YAX71	12312	2.3556 -	SFAM01.1	PPB	2/26/2022 14:52	SMPL
N1454-12	YAX72	10500	2.0001 -	SFAM01.1	PPB	2/26/2022 14:54	SMPL
N1454-13	YAX73	14188	2.7236 -	SFAM01.1	PPB	2/26/2022 14:57	SMPL
N1454-14	YAX74	4068	0.7384 -	SFAM01.1	PPB	2/26/2022 14:59	SMPL
CCV055	CCV055	24746	4.7947 -	SFAM01.1	PPB	2/26/2022 15:01	SMPL
CCB055	CCB055	259	-0.0088 -	SFAM01.1	PPB	2/26/2022 15:04	SMPL
N1454-15	YAX02	2240	0.3798 -	SFAM01.1	PPB	2/26/2022 15:06	SMPL
N1454-16	YAX65	19909	3.8459 -	SFAM01.1	PPB	2/26/2022 15:08	SMPL
N1454-17	YAX66	3145	0.5573 -	SFAM01.1	PPB	2/26/2022 15:10	SMPL
N1454-18	YAX67	2918	0.5128 -	SFAM01.1	PPB	2/26/2022 15:13	SMPL
N1454-19	YAX68	20382	3.9387 -	SFAM01.1	PPB	2/26/2022 15:15	SMPL
N1454-20	YAX69	2616	0.4536 -	SFAM01.1	PPB	2/26/2022 15:17	SMPL
N1454-21	YAX75	19379	3.7419 -	SFAM01.1	PPB	2/26/2022 15:19	SMPL
PB142638BL	PBS638	142	-0.0318 -	SFAM01.1	PPB	2/26/2022 15:22	SMPL
N1456-01	YAX32	3075	0.5436 -	SFAM01.1	PPB	2/26/2022 15:24	SMPL
N1456-02	YAX33	3236	0.5752 -	SFAM01.1	PPB	2/26/2022 15:26	SMPL
N1456-03	YAX34	2763	0.4824 -	SFAM01.1	PPB	2/26/2022 15:29	SMPL
N1456-04	YAX35	3019	0.5326 -	SFAM01.1	PPB	2/26/2022 15:31	SMPL
N1456-05	YAX36	4067	0.7382 -	SFAM01.1	PPB	2/26/2022 15:33	SMPL
N1456-06	YAX37	3068	0.5422 -	SFAM01.1	PPB	2/26/2022 15:35	SMPL
N1456-07	YAX38	3145	0.5573 -	SFAM01.1	PPB	2/26/2022 15:38	SMPL
N1456-08	YAX39	3823	0.6903 -	SFAM01.1	PPB	2/26/2022 15:40	SMPL
N1456-09	YAX40	3537	0.6342 -	SFAM01.1	PPB	2/26/2022 15:42	SMPL

LB118872
INSTRUMENT ID : CV1

N1456-10	YAX41	3085	0.5456 -	SFAM01.1	PPB	2/26/2022 15:45	SMPL
N1456-11	YAX42	3616	0.6497 -	SFAM01.1	PPB	2/26/2022 15:47	SMPL
N1456-12	YAX42D	3800	0.6858 -	SFAM01.1	PPB	2/26/2022 15:49	SMPL
CCV056	CCV056	25180	4.8799 -	SFAM01.1	PPB	2/26/2022 15:51	SMPL
CCB056	CCB056	274	-0.0059 -	SFAM01.1	PPB	2/26/2022 15:54	SMPL
N1456-13	YAX42S	15697	3.0196 -	SFAM01.1	PPB	2/26/2022 15:59	SMPL
N1456-14	YAX43	3470	0.6211 -	SFAM01.1	PPB	2/26/2022 16:01	SMPL
N1456-15	YAX44	3231	0.5742 -	SFAM01.1	PPB	2/26/2022 16:03	SMPL
N1456-16	YAX45	2594	0.4492 -	SFAM01.1	PPB	2/26/2022 16:05	SMPL
N1456-17	YAX46	3746	0.6752 -	SFAM01.1	PPB	2/26/2022 16:08	SMPL
N1456-18	YAX47	3019	0.5326 -	SFAM01.1	PPB	2/26/2022 16:10	SMPL
N1456-19	YAX56	14838	2.8511 -	SFAM01.1	PPB	2/26/2022 16:12	SMPL
N1456-20	YAX57	6735	1.2616 -	SFAM01.1	PPB	2/26/2022 16:15	SMPL
N1456-21	YAX58	3787	0.6833 -	SFAM01.1	PPB	2/26/2022 16:17	SMPL
N1456-22	YAX59	3342	0.596 -	SFAM01.1	PPB	2/26/2022 16:19	SMPL
PB142639BL	PBS639	140	-0.0322 -	SFAM01.1	PPB	2/26/2022 16:21	SMPL
N1458-01	YAX60	3630	0.6525 -	SFAM01.1	PPB	2/26/2022 16:24	SMPL
N1458-02	YAX61	3663	0.6589 -	SFAM01.1	PPB	2/26/2022 16:26	SMPL
N1458-03	YAX61D	3247	0.5773 -	SFAM01.1	PPB	2/26/2022 16:34	SMPL
N1458-04	YAX61S	14654	2.815 -	SFAM01.1	PPB	2/26/2022 16:36	SMPL
N1458-05	YAX78	1724	0.2786 -	SFAM01.1	PPB	2/26/2022 16:39	SMPL
PB142952BL	PBS952	398	0.0185 -	SFAM01.1	PPB	2/26/2022 16:41	SMPL
N1463-01	YAWW0	2646	0.4594 -	SFAM01.1	PPB	2/26/2022 16:43	SMPL
N1463-03	YAWW0S	15778	3.0355 -	SFAM01.1	PPB	2/26/2022 16:48	SMPL
CCV057	CCV057	24808	4.8069 -	SFAM01.1	PPB	2/26/2022 16:50	SMPL
CCB057	CCB057	264	-0.0078 -	SFAM01.1	PPB	2/26/2022 16:52	SMPL
N1463-02	YAWW0D	2686	0.4673 -	SFAM01.1	PPB	2/26/2022 16:55	SMPL
N1453-09DLX10	YAWX5	46069	8.9776 -	SFAM01.1	PPB	2/26/2022 16:57	SMPL
N1453-17DLX5	YAWY3	33314	6.4755 -	SFAM01.1	PPB	2/26/2022 16:59	SMPL
N1453-20DLX2	YAWY7	26279	5.0955 -	SFAM01.1	PPB	2/26/2022 17:02	SMPL
N1454-04DLX5	YAWZ0	19367	3.7396 -	SFAM01.1	PPB	2/26/2022 17:04	SMPL
N1454-05DLX2	YAWZ1	33402	6.4928 -	SFAM01.1	PPB	2/26/2022 17:06	SMPL
N1453-09DL2X25	YAWX5	18928	3.6534 -	SFAM01.1	PPB	2/26/2022 17:08	SMPL
CCV058	CCV058	24132	4.6743 -	SFAM01.1	PPB	2/26/2022 17:11	SMPL
CCB058	CCB058	270	-0.0067 -	SFAM01.1	PPB	2/26/2022 17:13	SMPL