



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/28/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:10
In Date: 02/27/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 02/28/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118876

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1631-01	DBRG1	1	1.00	1.00	2.00	2.00	100.0	T.B.
N1631-02	DBRG2	2	1.11	8.82	9.93	8.12	79.5	
N1631-03	DBRG2MS	3	1.11	8.82	9.93	8.12	79.5	
N1631-04	DBRG2MSD	4	1.11	8.82	9.93	8.12	79.5	
N1631-05	DBRG5	5	1.15	8.83	9.98	8.01	77.7	
N1631-06	DBRG6	6	1.12	8.72	9.84	7.63	74.7	
N1631-07	VHBLK001	7	1.00	1.00	2.00	2.00	100.0	VHBLK
N1631-08	DBRG7	8	1.00	1.00	2.00	2.00	100.0	T.B.
N1631-09	DBRG8	9	1.17	8.51	9.68	7.92	79.3	
N1631-10	DBRG9	10	1.14	8.47	9.61	7.95	80.4	
N1631-11	DBRH7	11	1.15	8.73	9.88	8.34	82.4	
N1631-12	DBRH8	12	1.13	8.59	9.72	7.93	79.2	
N1631-13	DBRH9	13	1.15	8.80	9.95	8.39	82.3	
N1631-14	DBRJ0	14	1.13	8.68	9.81	8.18	81.2	
N1631-15	DBRJ1	15	1.15	8.83	9.98	8.35	81.5	
N1631-16	DBRL6	16	1.12	8.67	9.79	8.12	80.7	
N1631-17	DBRL7	17	1.17	8.69	9.86	8.28	81.8	
N1631-18	DBRL8	18	1.11	8.87	9.98	8.66	85.1	
N1631-19	DBRL9	19	1.16	8.54	9.7	7.99	80.0	
N1631-20	DBRM0	20	1.16	8.81	9.97	8.65	85.0	
N1631-21	DBRM1	21	1.18	8.60	9.78	8.15	81.0	
N1631-22	DBRM2	22	1.17	8.74	9.91	8.26	81.1	
N1631-23	DBRM3	23	1.00	1.00	2.00	2.00	100.0	T.B.

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$