

LB118897


SFAM01.1

INSTRUMENT ID: CY1

Linear

 μ Abs.:

40414



A= 0.0000e+000

B= 2.4212e-004 *slope*C= 3.7337e-003 *y-intercept*

Rho= 0.9989974

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.00	0.000	0.027	0.027	97	0.000	97					-
0.05	0.050					185	0				15
0.20	0.200	0.230	0.030	933	0.0 %	933					-8
2.50	2.500	2.312	-0.188	9532	0.0 %	9532					1
5.00	5.000	5.063	0.063	20895	0.0 %	20895					4
7.50	7.500	7.780	0.280	32115	0.0 %	32115					-2
10.0	10.000	9.789	-0.211	40414	0.0 %	40414					

LB118877

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	97	-		0 SFAM01.1	PPB	2/27/2022 9:23	Std
	0.2 S01	933	-		0.2 SFAM01.1	PPB	2/27/2022 9:26	Std
	2.5 S02	9532	-		2.5 SFAM01.1	PPB	2/27/2022 9:29	Std
	5 S03	20895	-		5 SFAM01.1	PPB	2/27/2022 9:32	Std
	7.5 S04	32115	-		7.5 SFAM01.1	PPB	2/27/2022 9:34	Std
	10 S05	40414	-		10 SFAM01.1	PPB	2/27/2022 9:42	Std
ICV009	ICV009	16293	3.9487 -		SFAM01.1	PPB	2/27/2022 9:45	SMPL
ICB009	ICB009	22	0.0091 -		SFAM01.1	PPB	2/27/2022 9:47	SMPL
CCV059	CCV059	21263	5.152 -		SFAM01.1	PPB	2/27/2022 9:49	SMPL
CCB059	CCB059	42	0.0139 -		SFAM01.1	PPB	2/27/2022 9:51	SMPL
PB142961BL	PBW961	75	0.0219 -		SFAM01.1	PPB	2/27/2022 9:54	SMPL
N1454-22	YAX76	259	0.0664 -		SFAM01.1	PPB	2/27/2022 9:56	SMPL
PB142962BL	PBW962	79	0.0229 -		SFAM01.1	PPB	2/27/2022 9:58	SMPL
N1459-01	YAX77	6516	1.5814 -		SFAM01.1	PPB	2/27/2022 10:01	SMPL
N1459-02	YAX77D	5625	1.3657 -		SFAM01.1	PPB	2/27/2022 10:03	SMPL
N1459-03	YAX77S	9876	2.395 -		SFAM01.1	PPB	2/27/2022 10:11	SMPL
PB142963BL	PBW963	45	0.0146 -		SFAM01.1	PPB	2/27/2022 10:13	SMPL
N1619-01	YAY28	95	0.0267 -		SFAM01.1	PPB	2/27/2022 10:15	SMPL
N1619-02	YAY28D	95	0.0267 -		SFAM01.1	PPB	2/27/2022 10:18	SMPL
N1619-03	YAY28S	4541	1.1032 -		SFAM01.1	PPB	2/27/2022 10:20	SMPL
N1619-04	YAY29	83	0.0238 -		SFAM01.1	PPB	2/27/2022 10:22	SMPL
PB142964BL	PBW964	97	0.0272 -		SFAM01.1	PPB	2/27/2022 10:24	SMPL
N1620-01	YAY30	100	0.0279 -		SFAM01.1	PPB	2/27/2022 10:27	SMPL
N1620-02	YAY31	111	0.0306 -		SFAM01.1	PPB	2/27/2022 10:29	SMPL
N1620-03	YAY31D	115	0.0316 -		SFAM01.1	PPB	2/27/2022 10:31	SMPL
N1620-04	YAY31S	4813	1.1691 -		SFAM01.1	PPB	2/27/2022 10:36	SMPL
N1620-05	YAY34	103	0.0287 -		SFAM01.1	PPB	2/27/2022 10:38	SMPL
CCV060	CCV060	22380	5.4225 -		SFAM01.1	PPB	2/27/2022 10:41	SMPL
CCB060	CCB060	51	0.0161 -		SFAM01.1	PPB	2/27/2022 10:43	SMPL