




Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB118879
Analysis Date: 02/27/2022
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 02/27/2022
Extraction IN Time: 09:40
Extraction OUT Time: 10:25
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB118879BL	LB118879BL	WATER	1.3	1000	100	3.2869	3.2869	3.01	3.2870	3.2870	0.0001	0.1
2	LB118879BS	LB118879BS	WATER	1.3	1000	100	3.0143	3.0143	3.02	3.0315	3.0315	0.0172	17.2
3	LB118879BSD	LB118879BSD	WATER	1.3	1000	100	2.9637	2.9637	3.03	2.9811	2.9811	0.0174	17.4
4	N1645-07	LOD-MDL-WATER-01-QT1-2	WATER	1.3	1000	100	2.7814	2.7814	3.04	2.7836	2.7836	0.0022	2.2
5	N1645-08	LOQ-WATER-02-QT1-2022	WATER	1.3	1000	100	3.0147	3.0147	3.03	3.0189	3.0189	0.0042	4.2



Chemical Name	Chemical Lot #
HEXANE	W2885
pH Paper 0-14	M4909
Sodium Sulfate	EP2222
1:1 HCL	WP92935
Silica Gel	W2839
Sand	NA

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP92936
LCSWD	5.00 ML	WP92937
MS/MSD	NA	NA

Analytical Balance ID # : WC SC-6

0.0020 gram Balance:	<u>0.0021</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1	<u>11:31</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time1:	<u>11:00</u>		
Bal Check Time:	<u>10:00</u>		Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1	<u>12:05</u>
			Out Time1	<u>11:30</u>		

0.0020 gram Balance:	<u>0.0018</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>13:11</u>
1.0000 gram Balance:	<u>1.0004</u>	(0.9950-1.0050)	In Time2:	<u>12:37</u>		
Bal Check Time:	<u>14:20</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>13:50</u>
			Out Time2:	<u>13:10</u>		

118879

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POPLAR N1645 w WorkList ID : 157213 Department : Wet-Chemistry Date : 02-27-2022 09:17:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/14/2022	Water	N1645-07	Non-Polar Material	Conc H2SO4 to	CHEM02	QA Of	LOD-MDL-WATER-01-QT1-2022	02/23/2022	1664A
03/14/2022	Water	N1645-08	Non-Polar Material	Conc H2SO4 to	CHEM02	QA Of	LOQ-WATER-02-QT1-2022	02/23/2022	1664A

Date/Time 02/27/22 09:35
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: JD (WC)

Date/Time 02/27/22 12:40
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: JD (WC)