

LB118820


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

5522E



A= 0.0000e+000

B= 1.8405e-004

C= -5.7804e-002

Rho= 0.9998116

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.041	0.041	538	0.000	538					
0.2	0.200	0.225	0.025	1535	0.0 %	1535					13
2.5	2.500	2.511	0.011	13956	0.0 %	13956					0
5.0	5.000	4.902	-0.098	26945	0.0 %	26945					-2
7.5	7.500	7.415	-0.085	40600	0.0 %	40600					-1
10.0	10.000	10.107	0.107	55226	0.0 %	55226					1

LB118880 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 27 Feb 2022 12:46:08					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	538	-	0.0000	7471B	PPB	27 Feb 2022 12:53:16	S	Std
0.2 - 1	50.2	1535	-	0.2000	7471B	PPB	27 Feb 2022 12:55:34	S	Std
2.5 - 1	52.5	13956	-	2.5000	7471B	PPB	27 Feb 2022 12:57:51	S	Std
5.0 - 1	55	26945	-	5.0000	7471B	PPB	27 Feb 2022 13:00:08	S	Std
7.5 - 1	57.5	40600	-	7.5000	7471B	PPB	27 Feb 2022 13:02:25	S	Std
10.0 - 1	510	55226	-	10.0000	7471B	PPB	27 Feb 2022 13:12:22	S	Std
ICV10 - 1	ICV10	21573	3.9128	-	7471B	PPB	27 Feb 2022 13:15:12	U	SMPL
ICB10 - 1	ICB10	256	-0.0107	-	7471B	PPB	27 Feb 2022 13:17:27	U	SMPL
CCV61 - 1	CCV61	26784	4.8719	-	7471B	PPB	27 Feb 2022 13:19:42	U	SMPL
CCB61 - 1	CCB61	261	-0.0098	-	7471B	PPB	27 Feb 2022 13:21:57	U	SMPL
CRA - 1	CRA	1544	0.2264	-	7471B	PPB	27 Feb 2022 13:24:13	U	SMPL
HighStd - 1	HighStd	55967	10.2432	-	7471B	PPB	27 Feb 2022 13:26:29	U	SMPL
ChkStd - 1	ChkStd	37317	6.8106	-	7471B	PPB	27 Feb 2022 13:28:44	U	SMPL
PB142960BL - 1	PBS	381	0.0123	-	7471B	PPB	27 Feb 2022 13:31:02	U	SMPL
PB142960BS - 1	LCSS	21533	3.9055	-	7471B	PPB	27 Feb 2022 13:36:02	U	SMPL
N1659-01 - 1	PIT-11-SECTION-A	24766	4.5005	-	7471B	PPB	27 Feb 2022 13:38:18	U	SMPL
N1659-01DUP - 1	PIT-11-SECTION-ADUP	24631	4.4757	-	7471B	PPB	27 Feb 2022 13:42:52	U	SMPL
N1659-01MS - 1	PIT-11-SECTION-AMS	48399	8.8503	-	7471B	PPB	27 Feb 2022 13:45:09	U	SMPL
N1659-01MSD - 1	PIT-11-SECTION-AMSD	50229	9.1871	-	7471B	PPB	27 Feb 2022 13:47:26	U	SMPL
N1659-02 - 1	PIT-11-SECTION-B	25834	4.6971	-	7471B	PPB	27 Feb 2022 13:49:43	U	SMPL
CCV62 - 1	CCV62	26349	4.7919	-	7471B	PPB	27 Feb 2022 13:51:59	U	SMPL
CCB62 - 1	CCB62	250	-0.0118	-	7471B	PPB	27 Feb 2022 13:54:15	U	SMPL
N1678-01 - 1	OK-02-022522	1210	0.1649	-	7471B	PPB	27 Feb 2022 13:56:30	U	SMPL
N1679-01 - 1	RO-72-11944	1950	0.3011	-	7471B	PPB	27 Feb 2022 13:58:46	U	SMPL
N1679-04 - 1	RO-VNJ-221	2374	0.3791	-	7471B	PPB	27 Feb 2022 14:01:01	U	SMPL
N1681-01 - 1	TP-1-1-8FT	83691	15.3459	-	7471B	PPB	27 Feb 2022 14:03:17	U	SMPL
N1681-04 - 1	TP-2-1-8FT	58236	10.6608	-	7471B	PPB	27 Feb 2022 14:05:33	U	SMPL
N1681-07 - 1	TP-3-1-8FT	49808	9.1096	-	7471B	PPB	27 Feb 2022 14:07:51	U	SMPL
N1687-01 - 1	COMP-12-16-2021-DRUM	1141	0.1522	-	7471B	PPB	27 Feb 2022 14:10:10	U	SMPL
N1687-02 - 1	11-16-21-DRUM	1289	0.1794	-	7471B	PPB	27 Feb 2022 14:12:29	U	SMPL
N1687-03 - 1	12-29-21-DRUM	5291	0.9160	-	7471B	PPB	27 Feb 2022 14:14:47	U	SMPL
N1687-04 - 1	2-17-22-DRUM	2696	0.4384	-	7471B	PPB	27 Feb 2022 14:17:03	U	SMPL
CCV63 - 1	CCV63	26784	4.8719	-	7471B	PPB	27 Feb 2022 14:19:19	U	SMPL
CCB63 - 1	CCB63	325	0.0020	-	7471B	PPB	27 Feb 2022 14:21:35	U	SMPL
N1659-01LX5 - 1		5960	1.0392	-	7471B	PPB	27 Feb 2022 14:23:51	U	SMPL
N1659-01A - 1		47199	8.6294	-	7471B	PPB	27 Feb 2022 14:26:07	U	SMPL
N1681-01DLX5 - 1		17686	3.1974	-	7471B	PPB	27 Feb 2022 14:31:07	U	SMPL
N1681-04DLX2 - 1		29065	5.2918	-	7471B	PPB	27 Feb 2022 14:33:22	U	SMPL
CCV64 - 1	CCV64	27042	4.9194	-	7471B	PPB	27 Feb 2022 14:35:39	U	SMPL
CCB64 - 1	CCB64	272	-0.0077	-	7471B	PPB	27 Feb 2022 14:37:55	U	SMPL