

L8118891


INSTRUMENT ID: CV4

7470A

Linear

μ Abs.:

34821



A= 0.0000e+000

B= 2.8504e-004

C= 6.7825e-003

Rho= 0.9995563

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4/011
0.0	0.000	0.009	0.009	9	0.000	9					-6
0.2	0.200	0.213	0.013	724	0.0 %	724					-6
2.5	2.500	2.344	-0.156	8198	0.0 %	8198					4
5.0	5.000	5.208	0.208	18246	0.0 %	18246					0
7.5	7.500	7.494	-0.006	26267	0.0 %	26267					0
10.0	10.000	9.932	-0.068	34821	0.0 %	34821					-1

LB118891 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 28 Feb 2022 10:19:23				Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units				
0.0 - 1	50	9	-	0.0000	7470A	PPB	28 Feb 2022	10:23:15	S	Std
0.2 - 1	50.2	724	-	0.2000	7470A	PPB	28 Feb 2022	10:25:33	S	Std
2.5 - 1	52.5	8198	-	2.5000	7470A	PPB	28 Feb 2022	10:27:50	S	Std
5.0 - 1	55	18246	-	5.0000	7470A	PPB	28 Feb 2022	10:35:22	S	Std
7.5 - 1	57.5	26267	-	7.5000	7470A	PPB	28 Feb 2022	10:41:54	S	Std
10.0 - 1	51.0	34821	-	10.0000	7470A	PPB	28 Feb 2022	10:46:43	S	Std
ICV11 - 1	ICV11	14875	4.2468	-	7470A	PPB	28 Feb 2022	10:49:26	U	SMPL
ICB11 - 1	ICB11	-35	-0.0032	-	7470A	PPB	28 Feb 2022	10:51:41	U	SMPL
CCV65 - 1	CCV65	17835	5.0905	-	7470A	PPB	28 Feb 2022	10:53:56	U	SMPL
CCB65 - 1	CCB65	-37	-0.0038	-	7470A	PPB	28 Feb 2022	10:56:12	U	SMPL
CRA - 1	CRA	646	0.1909	-	7470A	PPB	28 Feb 2022	10:58:27	U	SMPL
HighStd - 1	HighStd	35000	9.9833	-	7470A	PPB	28 Feb 2022	11:00:42	U	SMPL
ChkStd - 1	ChkStd	25081	7.1559	-	7470A	PPB	28 Feb 2022	11:02:58	U	SMPL
PB142942BL - 1	PBW	-60	-0.0103	-	7470A	PPB	28 Feb 2022	11:05:14	U	SMPL
PB142942BS - 1	LCSW	15071	4.3027	-	7470A	PPB	28 Feb 2022	11:07:30	U	SMPL
N1664-02 - 1	TP-1	-28	-0.0012	-	7470A	PPB	28 Feb 2022	11:09:46	U	SMPL
N1664-02DUP - 1	TP-1DUP	5	0.0082	-	7470A	PPB	28 Feb 2022	11:12:03	U	SMPL
N1664-02MS - 1	TP-1MS	15678	4.4757	-	7470A	PPB	28 Feb 2022	11:17:51	U	SMPL
N1664-02MSD - 1	TP-1MSD	16480	4.7043	-	7470A	PPB	28 Feb 2022	11:20:08	U	SMPL
N1664-05 - 1	TP-2	-41	-0.0049	-	7470A	PPB	28 Feb 2022	11:22:23	U	SMPL
CCV66 - 1	CCV66	18843	5.3778	-	7470A	PPB	28 Feb 2022	11:24:39	U	SMPL
CCB66 - 1	CCB66	-34	-0.0029	-	7470A	PPB	28 Feb 2022	11:26:54	U	SMPL
N1664-08 - 1	TP-3	19	0.0122	-	7470A	PPB	28 Feb 2022	11:29:10	U	SMPL
N1664-11 - 1	TP-4	37	0.0173	-	7470A	PPB	28 Feb 2022	11:31:25	U	SMPL
N1664-14 - 1	TP-5	27	0.0145	-	7470A	PPB	28 Feb 2022	11:33:41	U	SMPL
N1664-17 - 1	TP-6	20	0.0125	-	7470A	PPB	28 Feb 2022	11:35:57	U	SMPL
N1664-20 - 1	TP-7	21	0.0128	-	7470A	PPB	28 Feb 2022	11:38:13	U	SMPL
N1664-23 - 1	TP-8	12	0.0102	-	7470A	PPB	28 Feb 2022	11:40:30	U	SMPL
N1664-26 - 1	TP-9	28	0.0148	-	7470A	PPB	28 Feb 2022	11:42:46	U	SMPL
N1664-29 - 1	TP-10	21	0.0128	-	7470A	PPB	28 Feb 2022	11:45:03	U	SMPL
N1664-32 - 1	TP-11	12	0.0102	-	7470A	PPB	28 Feb 2022	11:47:21	U	SMPL
PB142943BL - 1	PBW	-25	-0.0003	-	7470A	PPB	28 Feb 2022	11:49:37	U	SMPL
CCV67 - 1	CCV67	17311	4.9412	-	7470A	PPB	28 Feb 2022	11:54:39	U	SMPL
CCB67 - 1	CCB67	-41	-0.0049	-	7470A	PPB	28 Feb 2022	11:56:54	U	SMPL
PB142943BS - 1	LCSW	15003	4.2833	-	7470A	PPB	28 Feb 2022	11:59:10	U	SMPL
N1659-03 - 1	TP-3-MH	88	0.0319	-	7470A	PPB	28 Feb 2022	12:01:26	U	SMPL
N1659-03DUP - 1	TP-3-MHDUP	120	0.0410	-	7470A	PPB	28 Feb 2022	12:03:42	U	SMPL
N1659-03MS - 1	TP-3-MHMS	16447	4.6949	-	7470A	PPB	28 Feb 2022	12:05:57	U	SMPL
N1659-03MSD - 1	TP-3-MHMSD	16648	4.7522	-	7470A	PPB	28 Feb 2022	12:08:13	U	SMPL
N1661-01 - 1	MW-107	-34	-0.0029	-	7470A	PPB	28 Feb 2022	12:10:30	U	SMPL
N1661-02 - 1	MW-110	0	0.0068	-	7470A	PPB	28 Feb 2022	12:12:46	U	SMPL
N1661-03 - 1	MW-108	7	0.0088	-	7470A	PPB	28 Feb 2022	12:15:04	U	SMPL
N1661-04 - 1	MW-106	4	0.0079	-	7470A	PPB	28 Feb 2022	12:17:21	U	SMPL
N1661-05 - 1	MW-107	4	0.0079	-	7470A	PPB	28 Feb 2022	12:19:37	U	SMPL
CCV68 - 1	CCV68	17557	5.0113	-	7470A	PPB	28 Feb 2022	12:23:55	U	SMPL
CCB68 - 1	CCB68	-17	0.0019	-	7470A	PPB	28 Feb 2022	12:26:10	U	SMPL
N1661-06 - 1	MW-110	0	0.0068	-	7470A	PPB	28 Feb 2022	12:28:26	U	SMPL
N1661-07 - 1	MW-108	12	0.0102	-	7470A	PPB	28 Feb 2022	12:30:42	U	SMPL
N1661-08 - 1	MW-106	23	0.0133	-	7470A	PPB	28 Feb 2022	12:32:58	U	SMPL
N1668-01 - 1	PURGED-WATER	90	0.0324	-	7470A	PPB	28 Feb 2022	12:35:13	U	SMPL
PB142918TB - 1	PB142918TB	14	0.0108	-	7470A	PPB	28 Feb 2022	12:37:31	U	SMPL
N1664-02LX5 - 1		-6	0.0051	-	7470A	PPB	28 Feb 2022	12:39:47	U	SMPL
N1664-02A - 1		15417	4.4013	-	7470A	PPB	28 Feb 2022	12:42:04	U	SMPL
N1659-03LX5 - 1		-22	0.0005	-	7470A	PPB	28 Feb 2022	12:44:21	U	SMPL
N1659-03A - 1		15571	4.4452	-	7470A	PPB	28 Feb 2022	12:46:38	U	SMPL
CCV69 - 1	CCV69	17605	5.0250	-	7470A	PPB	28 Feb 2022	12:48:54	U	SMPL
CCB69 - 1	CCB69	-28	-0.0012	-	7470A	PPB	28 Feb 2022	12:51:10	U	SMPL