



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB118894
Thermometer ID: flashpoint

Reviewed By: ketankumar
Supervisor Review By: apatel
Ambient Barometric Pressure (mmHg): 765.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	03/01/2022	09:35
2	N1638-01		1	10	100.00	>212.0	>212.0	03/01/2022	09:50
3	N1638-01DUP		1	10	100.00	>212.0	>212.0	03/01/2022	10:10
4	N1638-02		1	15	100.00	>212.0	>212.0	03/01/2022	10:32
5	N1642-07		1	15	100.00	>212.0	>212.0	03/01/2022	10:55
6	N1642-08		1	12	100.00	>212.0	>212.0	03/01/2022	11:15
7	N1647-01		1	14	100.00	>212.0	>212.0	03/01/2022	11:35
8	N1647-02		1	13	100.00	>212.0	>212.0	03/01/2022	11:58
9	N1647-03		1	15	100.00	>212.0	>212.0	03/01/2022	12:20
10	N1659-03		1	12	100.00	>212.0	>212.0	03/01/2022	12:45
11	N1668-01		1	14	100.00	>212.0	>212.0	03/01/2022	13:10

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

EB 11 8899

WorkList Name : fp22522 WorkList ID : 157168 Department : Wet-Chemistry Date : 02-25-2022 09:15:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/10/2022	Water	N1659-03	Flash Point	Cool 4 deg C	TULL02	P11	TP-3-MH	02/24/2022	1010B
03/02/2022	Water	N1638-01	Flash Point	Cool 4 deg C	PSEG03	M11	30960	02/23/2022	1010B
03/02/2022	Water	N1638-02	Flash Point	Cool 4 deg C	PSEG03	M11	40473	02/23/2022	1010B
03/02/2022	Water	N1642-07	Flash Point	Cool 4 deg C	PSEG03	M11	31453-58	02/23/2022	1010B
03/02/2022	Water	N1642-08	Flash Point	Cool 4 deg C	PSEG03	M11	31409	02/23/2022	1010B
03/02/2022	Water	N1647-01	Flash Point	Cool 4 deg C	PSEG03	P11	31382	02/23/2022	1010B
03/02/2022	Water	N1647-02	Flash Point	Cool 4 deg C	PSEG03	P11	31415	02/23/2022	1010B
03/02/2022	Water	N1647-03	Flash Point	Cool 4 deg C	PSEG03	P11	31416	02/23/2022	1010B
03/03/2022	Water	N1668-01	Flash Point	Cool 4 deg C	PSEG03	M11	PURGED-WATER	02/24/2022	1010B

Date/Time 03/01/22 09:15
 Raw Sample Received by: RS (WC)
 Raw Sample Relinquished by: RS (WC)

Date/Time 03/01/22 13:15
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: RS (WC)