



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB118897

Reagent/Standard	Lot/Log #
COD Digestion Vials Low Level 0-150Mg/L	W2829
COD ICV-LCS std, 50ppm	WP95509
COD CCV std, 50ppm	WP95508
COD LOD std, 5ppm	WP95510
COD LOQ std, 10.0PPM	WP95511
COD ICV-LCS std, 50ppm	WP95371
COD calibration std. 100 ppm	WP95373
COD calibration std. 10 ppm	WP95375
COD calibration std. 150 ppm	WP95372
COD calibration std. 50 ppm	WP95374
COD calibration std. 0 ppm	WP95376

Temp In(C): 150	Date In: 02/25/2022	Time In: 11:30
Temp Out(C): 150	Date Out: 02/25/2022	Time Out: 13:30

Intercept: -0.4433

Slope: 1.0136

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	0.437		02/18/2022	13:20
2	CAL2	10	1	Water	9.000	9.317	6.8	02/18/2022	13:20
3	CAL3	50	1	Water	51.000	50.753	-1.5	02/18/2022	13:21
4	CAL4	100	1	Water	100.000	99.096	0.9	02/18/2022	13:21
5	CAL5	150	1	Water	152.000	150.398	-0.3	02/18/2022	13:22



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB118897

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	49.000	48.780	02/18/2022	13:22
2	ICB		NA	NA	1	Water	0.000	0.437	02/18/2022	13:23
3	CCV1	50	NA	NA	1	Water	50.000	49.766	02/25/2022	15:30
4	CCB1		NA	NA	1	Water	1.000	1.424	02/25/2022	15:30
5	LB118897BL		1	50	1	Solid	0.000	0.437	02/25/2022	15:31
6	LB118897BS	50	1	50	1	Solid	49.000	48.780	02/25/2022	15:32
7	N1645-01		1	50	1	Solid	5.000	5.370	02/25/2022	15:32
8	N1645-02		1	50	1	Solid	9.000	9.317	02/25/2022	15:33
9	CCV2	50	NA	NA	1	Water	49.000	48.780	02/25/2022	15:33
10	CCB2		NA	NA	1	Water	0.000	0.437	02/25/2022	15:34