


LB118909

7470A

INSTRUMENT ID: CV1

 μ Abs.:

40368



Linear

A= 0.0000e+000

B= 2.4986e-004

C= 5.2664e-002

Rho= 0.9996961

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/d
0.0	0.000	0.065	0.065	50	0.000	50					-
0.2	0.200	0.220	0.020	669	0.0 %	669					10
2.5	2.500	2.480	-0.020	9713	0.0 %	9713					-1
5.0	5.000	4.946	-0.054	19586	0.0 %	19586					-1
7.5	7.500	7.350	-0.150	29207	0.0 %	29207					-2
10.0	10.000	10.139	0.139	40368	0.0 %	40368					1

LB118907 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Feb 2022 14:40:57

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	50	-	0.0000	7470A	PPB	28 Feb 2022 14:45:09	S	Std
0.2 - 1	50.2	669	-	0.2000	7470A	PPB	28 Feb 2022 14:47:27	S	Std
2.5 - 1	52.5	9713	-	2.5000	7470A	PPB	28 Feb 2022 14:49:45	S	Std
5.0 - 1	55	19586	-	5.0000	7470A	PPB	28 Feb 2022 14:52:02	S	Std
7.5 - 1	57.5	29207	-	7.5000	7470A	PPB	28 Feb 2022 14:54:18	S	Std
10.0 - 1	510	40368	-	10.0000	7470A	PPB	28 Feb 2022 15:01:52	S	Std
ICV12 - 1	ICV12	14239	3.6104	-	7470A	PPB	28 Feb 2022 15:05:02	U	SMPL
ICB12 - 1	ICB12	120	0.0826	-	7470A	PPB	28 Feb 2022 15:07:17	U	SMPL
CCV70 - 1	CCV70	18534	4.6835	-	7470A	PPB	28 Feb 2022 15:09:32	U	SMPL
CCB70 - 1	CCB70	134	0.0861	-	7470A	PPB	28 Feb 2022 15:11:48	U	SMPL
CRA - 1	CRA	671	0.2203	-	7470A	PPB	28 Feb 2022 15:14:03	U	SMPL
HighStd - 1	HighStd	40948	10.2838	-	7470A	PPB	28 Feb 2022 15:16:18	U	SMPL
ChkStd - 1	ChkStd	26612	6.7019	-	7470A	PPB	28 Feb 2022 15:18:34	U	SMPL
PB142970BL - 1	PBW	-33	0.0444	-	7470A	PPB	28 Feb 2022 15:20:50	U	SMPL
PB142970BS - 1	LCSW	14277	3.6199	-	7470A	PPB	28 Feb 2022 15:23:06	U	SMPL
N1626-18 - 1	CC-B1	2	0.0532	-	7470A	PPB	28 Feb 2022 15:25:22	U	SMPL
N1626-18DUP - 1	CC-B1DUP	30	0.0602	-	7470A	PPB	28 Feb 2022 15:27:39	U	SMPL
N1626-18MS - 1	CC-B1MS	13397	3.4000	-	7470A	PPB	28 Feb 2022 15:29:56	U	SMPL
N1626-18MSD - 1	CC-B1MSD	14621	3.7058	-	7470A	PPB	28 Feb 2022 15:32:13	U	SMPL
N1626-19 - 1	CC-B2	32	0.0607	-	7470A	PPB	28 Feb 2022 15:34:29	U	SMPL
CCV71 - 1	CCV71	18575	4.6938	-	7470A	PPB	28 Feb 2022 15:36:44	U	SMPL
CCB71 - 1	CCB71	143	0.0884	-	7470A	PPB	28 Feb 2022 15:39:00	U	SMPL
N1626-20 - 1	CC-B4	15	0.0564	-	7470A	PPB	28 Feb 2022 15:41:15	U	SMPL
N1626-21 - 1	CC-B8	174	0.0961	-	7470A	PPB	28 Feb 2022 15:43:31	U	SMPL
PB142970TB - 1	PB142951TB	33	0.0609	-	7470A	PPB	28 Feb 2022 15:45:46	U	SMPL
N1626-18LX5 - 1		161	0.0929	-	7470A	PPB	28 Feb 2022 15:48:02	U	SMPL
N1626-18A - 1		15149	3.8377	-	7470A	PPB	28 Feb 2022 15:50:19	U	SMPL
CCV72 - 1	CCV72	18678	4.7195	-	7470A	PPB	28 Feb 2022 15:52:36	U	SMPL
CCB72 - 1	CCB72	130	0.0851	-	7470A	PPB	28 Feb 2022 15:54:53	U	SMPL