



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118911

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/28/2022	16:15
2	CAL2	1	Water	NA	NA	20.6	7.00	02/28/2022	20:06
3	CAL3	1	Water	NA	NA	20.6	10.03	02/28/2022	16:17
4	ICV	1	Water	NA	NA	20.7	6.99	02/28/2022	16:19
5	CCV1	1	Water	NA	NA	20.7	2.01	02/28/2022	16:20
6	N1712-03	1	Water	NA	NA	20.1	6.26	02/28/2022	16:35
7	N1712-03DUP	1	Water	NA	NA	20.2	6.27	02/28/2022	16:28
8	CCV2	1	Water	NA	NA	20.7	12.03	02/28/2022	16:39

WORKLIST(Hardcopy Internal Chain)

118911

WorkList Name : PH N1712 WorkList ID : 157263 Department : Wet-Chemistry Date : 02-28-2022 16:09:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/28/2022	Water	N1712-03	pH		Cool 4 deg C	P11	31404	02/28/2022	9040C

Date/Time 02/28/22 16:10
 Raw Sample Received by: JP WCI
 Raw Sample Relinquished by: JP WCI

Date/Time 02/28/22 17:30
 Raw Sample Received by: JP WCI
 Raw Sample Relinquished by: JP WCI