



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118912

Slope : 99.2

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	02/28/2022	16:15
2	CAL2	1	Water	NA	NA	20.6	7.00	02/28/2022	16:16
3	CAL3	1	Water	NA	NA	20.6	10.03	02/28/2022	16:17
4	ICV	1	Water	NA	NA	20.7	6.99	02/28/2022	16:19
5	CCV1	1	Water	NA	NA	20.7	2.01	02/28/2022	16:20
6	N1703-01	1	Solid	20.03	20	20.8	9.30	02/28/2022	16:22
7	N1713-01	1	Solid	20.04	20	20.9	6.08	02/28/2022	16:25
8	N1713-04	1	Solid	20.03	20	20.3	5.55	02/28/2022	16:28
9	N1713-04DUP	1	Solid	20.04	20	20.4	5.56	02/28/2022	16:29
10	CCV2	1	Water	NA	NA	20.4	12.03	02/28/2022	16:30

WORKLIST(Hardcopy Internal Chain)

WB 118912

WorkList Name : CORROSIONITY N1713 WorkList ID : 157262 Department : Wet-Chemistry Date : 02-28-2022 15:31:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/28/2022	Solid	N1703-01	Corrosivity	Cool 4 deg C	PSEG03	N11	OILY-STONE-DRUMS	02/28/2022	9045D
02/28/2022	Solid	N1713-01	Corrosivity	Cool 4 deg C	PSEG03	P11	DRUM-COMP	02/28/2022	9045D
02/28/2022	Solid	N1713-04	Corrosivity	Cool 4 deg C	PSEG03	P11	VNJ-235	02/28/2022	9045D

Date/Time 02/28/22 16:10
 Raw Sample Received by: JH JWC
 Raw Sample Relinquished by: CP SM

Date/Time 02/28/22 17:30
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JH JWC