

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

2/28/2022 16:28

Reviewed by : RL Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.977	0.0	0.108	
ICB1	0.010	0.0	0.007	
CCV1	0.985	0.0	0.109	
CCB1	0.029	0.0	0.009	
PB142957BL	0.020	0.0	0.008	
PB142957BS	0.979	0.0	0.108	
N1645-02	0.056	0.0	0.011	
N1645-01	0.029	0.0	0.009	
PB142959BL	0.015	0.0	0.007	
PB142959BS	0.975	0.0	0.108	
N1645-07	0.027	0.0	0.008	
N1645-08	0.048	0.0	0.011	
CCV2	0.973	0.0	0.108	
CCB2	0.014	0.0	0.007	

N 14
Mean 0.367
SD 0.4726
CV% 128.88

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 28 14:31:47 2022

Mon Feb 28 16:28:36 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0037	mg/l	2/28/2022 14:49:10	
0.05PPM	A	Phenolics-	P	0.057	mg/l	2/28/2022 14:49:11	
0.1PPM	A	Phenolics-	P	0.1058	mg/l	2/28/2022 14:49:12	
0.25PPM	A	Phenolics-	P	0.2551	mg/l	2/28/2022 14:49:13	
0.50PPM	A	Phenolics-	P	0.4945	mg/l	2/28/2022 14:49:14	
1.0PPM	A	Phenolics-	P	0.9672	mg/l	2/28/2022 14:49:15	
2.0PPM	A	Phenolics-	P	2.0166	mg/l	2/28/2022 14:49:16	
ICV1	S	Phenolics-	P	0.9768	mg/l	2/28/2022 15:49:17	
ICB1	S	Phenolics-	P	0.01	mg/l	2/28/2022 15:49:18	
CCV1	S	Phenolics-	P	0.9848	mg/l	2/28/2022 15:49:21	
CCB1	S	Phenolics-	P	0.0287	mg/l	2/28/2022 15:49:22	
PB142957BL	S	Phenolics-	P	0.02	mg/l	2/28/2022 15:49:24	
PB142957BS	S	Phenolics-	P	0.9786	mg/l	2/28/2022 15:49:26	
N1645-02	S	Phenolics-	P	0.056	mg/l	2/28/2022 15:59:20	
N1645-01	S	Phenolics-	P	0.029	mg/l	2/28/2022 16:24:43	
PB142959BL	S	Phenolics-	P	0.0145	mg/l	2/28/2022 16:24:45	
PB142959BS	S	Phenolics-	P	0.9748	mg/l	2/28/2022 16:24:46	
N1645-07	S	Phenolics-	P	0.0266	mg/l	2/28/2022 16:24:47	
N1645-08	S	Phenolics-	P	0.048	mg/l	2/28/2022 16:24:48	
CCV2	S	Phenolics-	P	0.9726	mg/l	2/28/2022 16:28:35	
CCB2	S	Phenolics-	P	0.0135	mg/l	2/28/2022 16:28:36	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

2/28/2022 14:49

Reviewed by : 12 Instrument ID : Konelab


Test Phenolics-

Accepted 2/28/2022 14:49

Factor 9.531
Bias 0.006

Coeff. of det. 0.999523

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.006	0.0037	0.0000	-
2	Phenol-2	0.012	0.0570	0.0500	14.0
3	Phenol-2	0.017	0.1058	0.1000	5.8
4	Phenol-2	0.032	0.2551	0.2500	2.0
5	Phenol-2	0.057	0.4945	0.5000	-1.1
6	Phenol-2	0.107	0.9672	1.0000	-3.3
7	Phenol-2	0.217	2.0166	2.0000	0.8

12 02/28/22