

LB118915

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

2/28/2022 16:57

Test: Phenolics-

Sample Id	<i>mp/c</i> Result	Dil. 1 +	Response	Errors
ICV1	0.981	0.0	0.109	
ICB1	0.016	0.0	0.007	
CCV1	0.985	0.0	0.109	
CCB1	0.015	0.0	0.007	
PB142958BL	0.014	0.0	0.007	
PB142958BS	0.981	0.0	0.109	
N1645-10	0.028	0.0	0.009	
N1645-11	0.049	0.0	0.011	
CCV2	0.983	0.0	0.109	
CCB2	0.014	0.0	0.007	

N	10
Mean	0.407
SD	0.4959
CV%	121.96

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 28 16:50:38 2022

Mon Feb 28 16:57:31 2022

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0037	mg/l	2/28/2022 14:49:10	
0.05PPM	A	Phenolics-	P	0.057	mg/l	2/28/2022 14:49:11	
0.1PPM	A	Phenolics-	P	0.1058	mg/l	2/28/2022 14:49:12	
0.25PPM	A	Phenolics-	P	0.2551	mg/l	2/28/2022 14:49:13	
0.50PPM	A	Phenolics-	P	0.4945	mg/l	2/28/2022 14:49:14	
1.0PPM	A	Phenolics-	P	0.9672	mg/l	2/28/2022 14:49:15	
2.0PPM	A	Phenolics-	P	2.0166	mg/l	2/28/2022 14:49:16	
ICV1	S	Phenolics-	P	0.9812	mg/l	2/28/2022 16:50:39	
ICB1	S	Phenolics-	P	0.0157	mg/l	2/28/2022 16:50:41	
CCV1	S	Phenolics-	P	0.9855	mg/l	2/28/2022 16:50:43	
CCB1	S	Phenolics-	P	0.0147	mg/l	2/28/2022 16:50:45	
PB142958BL	S	Phenolics-	P	0.0144	mg/l	2/28/2022 16:50:46	
PB142958BS	S	Phenolics-	P	0.9809	mg/l	2/28/2022 16:50:48	
N1645-10	S	Phenolics-	P	0.028	mg/l	2/28/2022 16:57:24	
N1645-11	S	Phenolics-	P	0.0486	mg/l	2/28/2022 16:57:26	
CCV2	S	Phenolics-	P	0.9828	mg/l	2/28/2022 16:57:29	
CCB2	S	Phenolics-	P	0.014	mg/l	2/28/2022 16:57:31	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

2/28/2022 14:49


Test Phenolics-

Accepted 2/28/2022 14:49

Factor 9.531
Bias 0.006

Coeff. of det. 0.999523

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.006	0.0037	0.0000	-
2	Phenol-2	0.012	0.0570	0.0500	14.0
3	Phenol-2	0.017	0.1058	0.1000	5.8
4	Phenol-2	0.032	0.2551	0.2500	2.0
5	Phenol-2	0.057	0.4945	0.5000	-1.1
6	Phenol-2	0.107	0.9672	1.0000	-3.3
7	Phenol-2	0.217	2.0166	2.0000	0.8

12 02/28/22