


7471B

LB118924
INSTRUMENT ID: CV2

Linear

 μ Abs.:

54968



A= 0.0000e+000

B= 1.8212e-004

C= 3.9268e-002

Rho= 0.9998219

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/aD
0.0	0.000	0.034	0.034	-29	0.000	-29					7
0.2	0.200	0.215	0.015	964	0.0 %	964					1
2.5	2.500	2.535	0.035	13704	0.0 %	13704					3
5.0	5.000	4.846	-0.154	26395	0.0 %	26395					0
7.5	7.500	7.519	0.019	41072	0.0 %	41072					1
10.0	10.000	10.050	0.050	54968	0.0 %	54968					

LB118924 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Mar 2022 10:47:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-29	-	0.0000	7471B	PPB	01 Mar 2022 10:57:01	S	Std
0.2 - 1	50.2	964	-	0.2000	7471B	PPB	01 Mar 2022 10:59:18	S	Std
2.5 - 1	52.5	13704	-	2.5000	7471B	PPB	01 Mar 2022 11:01:35	S	Std
5.0 - 1	55	26395	-	5.0000	7471B	PPB	01 Mar 2022 11:03:53	S	Std
7.5 - 1	57.5	41072	-	7.5000	7471B	PPB	01 Mar 2022 11:06:10	S	Std
10.0 - 1	510	54968	-	10.0000	7471B	PPB	01 Mar 2022 11:08:27	S	Std
ICV13 - 1	ICV13	21198	3.8999	-	7471B	PPB	01 Mar 2022 11:11:16	U	SMPL
ICB13 - 1	ICB13	-22	0.0353	-	7471B	PPB	01 Mar 2022 11:13:31	U	SMPL
CCV73 - 1	CCV73	26020	4.7781	-	7471B	PPB	01 Mar 2022 11:15:46	U	SMPL
CCB73 - 1	CCB73	-34	0.0331	-	7471B	PPB	01 Mar 2022 11:18:02	U	SMPL
CRA - 1	CRA	958	0.2137	-	7471B	PPB	01 Mar 2022 11:20:17	U	SMPL
HighStd - 1	HighStd	50095	9.1627	-	7471B	PPB	01 Mar 2022 11:22:32	U	SMPL
ChkStd - 1	ChkStd	36979	6.7740	-	7471B	PPB	01 Mar 2022 11:24:49	U	SMPL
PB143001BL - 1	PBS	-70	0.0265	-	7471B	PPB	01 Mar 2022 11:27:08	U	SMPL
PB143001BS - 1	LCSS	21739	3.9985	-	7471B	PPB	01 Mar 2022 11:29:25	U	SMPL
N1704-01 - 1	2197-A	49664	9.0842	-	7471B	PPB	01 Mar 2022 11:31:41	U	SMPL
N1705-01 - 1	2171-B	17370	3.2028	-	7471B	PPB	01 Mar 2022 11:33:58	U	SMPL
N1706-01 - 1	2181-C	2155	0.4317	-	7471B	PPB	01 Mar 2022 11:36:16	U	SMPL
N1711-01 - 1	NB-07-022822	2924	0.5718	-	7471B	PPB	01 Mar 2022 11:38:33	U	SMPL
N1711-01DUP - 1	NB-07-022822DUP	2951	0.5767	-	7471B	PPB	01 Mar 2022 11:40:49	U	SMPL
CCV74 - 1	CCV74	25966	4.7683	-	7471B	PPB	01 Mar 2022 11:43:04	U	SMPL
CCB74 - 1	CCB74	-51	0.0300	-	7471B	PPB	01 Mar 2022 11:45:19	U	SMPL
N1711-01MS - 1	NB-07-022822MS	24437	4.4898	-	7471B	PPB	01 Mar 2022 11:47:35	U	SMPL
N1711-01MSD - 1	NB-07-022822MSD	23719	4.3591	-	7471B	PPB	01 Mar 2022 11:49:51	U	SMPL
N1712-01 - 1	31421-22-23	555	0.1403	-	7471B	PPB	01 Mar 2022 11:52:07	U	SMPL
N1713-03 - 1	VNJ-235	628	0.1536	-	7471B	PPB	01 Mar 2022 11:54:23	U	SMPL
N1711-01LX5 - 1		542	0.1380	-	7471B	PPB	01 Mar 2022 11:56:39	U	SMPL
N1711-01A - 1		24228	4.4518	-	7471B	PPB	01 Mar 2022 11:58:56	U	SMPL
CCV75 - 1	CCV75	25853	4.7477	-	7471B	PPB	01 Mar 2022 12:01:12	U	SMPL
CCB75 - 1	CCB75	-45	0.0311	-	7471B	PPB	01 Mar 2022 12:03:30	U	SMPL