

L13 118935

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

3/1/2022 15:13

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.992	0.0	0.104	
ICB1	0.003	0.0	0.005	
CCV1	0.999	0.0	0.105	
CCB1	0.002	0.0	0.005	
PB142995BL	0.002	0.0	0.005	
PB142995BS	0.991	0.0	0.104	
N1645-12	0.028	0.0	0.008	
CCV2	0.997	0.0	0.105	
CCB2	0.002	0.0	0.005	

N	9
Mean	0.446
SD	0.5205
CV%	116.64

Aquakem v. 7.2AQ1

Results from time period:

Tue Mar 01 15:06:57 2022

Tue Mar 01 15:12:52 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0037	mg/l	2/28/2022 14:49:10	
0.05PPM	A	Phenolics-	P	0.057	mg/l	2/28/2022 14:49:11	
0.1PPM	A	Phenolics-	P	0.1058	mg/l	2/28/2022 14:49:12	
0.25PPM	A	Phenolics-	P	0.2551	mg/l	2/28/2022 14:49:13	
0.50PPM	A	Phenolics-	P	0.4945	mg/l	2/28/2022 14:49:14	
1.0PPM	A	Phenolics-	P	0.9672	mg/l	2/28/2022 14:49:15	
2.0PPM	A	Phenolics-	P	2.0166	mg/l	2/28/2022 14:49:16	
ICV1	S	Phenolics-	P	0.9915	mg/l	3/1/2022 15:06:57	
ICB1	S	Phenolics-	P	0.0026	mg/l	3/1/2022 15:06:59	
CCV1	S	Phenolics-	P	0.999	mg/l	3/1/2022 15:07:01	
CCB1	S	Phenolics-	P	0.0021	mg/l	3/1/2022 15:07:03	
PB142995BL	S	Phenolics-	P	0.0024	mg/l	3/1/2022 15:07:05	
PB142995BS	S	Phenolics-	P	0.9913	mg/l	3/1/2022 15:07:07	
N1645-12	S	Phenolics-	P	0.028	mg/l	3/1/2022 15:12:47	
CCV2	S	Phenolics-	P	0.9974	mg/l	3/1/2022 15:12:50	
CCB2	S	Phenolics-	P	0.0017	mg/l	3/1/2022 15:12:51	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

3/1/2022 14:06

Test Phenolics-


Accepted 3/1/2022 14:01

Factor 9.958

Bias 0.005

Coeff. of det. 0.999739

Errors



	Calibrator	Response	Calc. con.	Conc.	Re Errors
1	0.00PPM	0.005	0.0015	0.0000	-
2	Phenol-2	0.010	0.0548	0.0500	9.6
3	Phenol-2	0.015	0.1052	0.1000	5.2
4	Phenol-2	0.030	0.2551	0.2500	2.0
5	Phenol-2	0.055	0.4953	0.5000	-0.9
6	Phenol-2	0.103	0.9760	1.0000	-2.4
7	Phenol-2	0.207	2.0121	2.0000	0.6

12 3/1/22