

LB118946


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40453



A= 0.0000e+000

B= 2.4748e-004

C= 5.9129e-002

Rho= 0.9999026

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.056	0.056	-11	0.000	-11					9
0.2	0.200	0.219	0.019	647	0.0 %	647					-2
2.5	2.500	2.457	-0.043	9688	0.0 %	9688					-1
5.0	5.000	4.938	-0.062	19714	0.0 %	19714					-1
7.5	7.500	7.459	-0.041	29901	0.0 %	29901					-1
10.0	10.000	10.071	0.071	40453	0.0 %	40453					1

LB118946 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Mar 2022 14:19:54

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-11	-	0.0000	7470A	PPB	02 Mar 2022 14:21:34	S	Std
0.2 - 1	50.2	647	-	0.2000	7470A	PPB	02 Mar 2022 14:23:52	S	Std
2.5 - 1	52.5	9688	-	2.5000	7470A	PPB	02 Mar 2022 14:26:09	S	Std
5.0 - 1	55	19714	-	5.0000	7470A	PPB	02 Mar 2022 14:28:26	S	Std
7.5 - 1	57.5	29901	-	7.5000	7470A	PPB	02 Mar 2022 14:30:43	S	Std
10.0 - 1	59	40453	-	10.0000	7470A	PPB	02 Mar 2022 14:32:59	S	Std
ICV15 - 1	ICV15	15235	3.8295	-	7470A	PPB	02 Mar 2022 14:39:00	U	SMPL
ICB15 - 1	ICB15	-36	0.0502	-	7470A	PPB	02 Mar 2022 14:41:15	U	SMPL
CCV81 - 1	CCV81	18493	4.6358	-	7470A	PPB	02 Mar 2022 14:43:30	U	SMPL
CCB81 - 1	CCB81	-50	0.0468	-	7470A	PPB	02 Mar 2022 14:45:47	U	SMPL
CRA - 1	CRA	603	0.2084	-	7470A	PPB	02 Mar 2022 14:48:02	U	SMPL
HighStd - 1	HighStd	39921	9.9389	-	7470A	PPB	02 Mar 2022 14:50:17	U	SMPL
ChkStd - 1	ChkStd	26974	6.7347	-	7470A	PPB	02 Mar 2022 14:52:32	U	SMPL
PB143014BL - 1	PBW	-62	0.0438	-	7470A	PPB	02 Mar 2022 14:54:49	U	SMPL
PB143014BS - 1	LCSW	16128	4.0505	-	7470A	PPB	02 Mar 2022 14:57:05	U	SMPL
N1713-04 - 1	VNU-235	-48	0.0472	-	7470A	PPB	02 Mar 2022 14:59:22	U	SMPL
N1713-04DUP - 1	VNU-235DUP	1	0.0594	-	7470A	PPB	02 Mar 2022 15:01:39	U	SMPL
N1713-04MS - 1	VNU-235MS	14208	3.5754	-	7470A	PPB	02 Mar 2022 15:03:56	U	SMPL
N1713-04MSD - 1	VNU-235MSD	14566	3.6640	-	7470A	PPB	02 Mar 2022 15:06:14	U	SMPL
PB143022BL - 1	PBW	-37	0.0500	-	7470A	PPB	02 Mar 2022 15:08:29	U	SMPL
CCV82 - 1	CCV82	18844	4.7227	-	7470A	PPB	02 Mar 2022 15:10:44	U	SMPL
CCB82 - 1	CCB82	-36	0.0502	-	7470A	PPB	02 Mar 2022 15:13:00	U	SMPL
PB143022BS - 1	LCSW	15067	3.7879	-	7470A	PPB	02 Mar 2022 15:15:15	U	SMPL
N1701-01 - 1	20220213-096-21-01-ELUTRIATE	27	0.0658	-	7470A	PPB	02 Mar 2022 15:17:31	U	SMPL
N1701-01DUP - 1	20220213-096-21-01-ELUTRIATEDUP	48	0.0710	-	7470A	PPB	02 Mar 2022 15:19:46	U	SMPL
N1701-01MS - 1	20220213-096-21-01-ELUTRIATEMS	16089	4.0409	-	7470A	PPB	02 Mar 2022 15:22:02	U	SMPL
N1701-01MSD - 1	20220213-096-21-01-ELUTRIATEMSD	14657	3.6865	-	7470A	PPB	02 Mar 2022 15:24:18	U	SMPL
N1701-02 - 1	20220214-096-21-02-ELUTRIATE	10	0.0616	-	7470A	PPB	02 Mar 2022 15:26:35	U	SMPL
PB142998TB - 1	PB142998TB	-44	0.0482	-	7470A	PPB	02 Mar 2022 15:28:53	U	SMPL
CCV83 - 1	CCV83	20080	5.0286	-	7470A	PPB	02 Mar 2022 15:36:08	U	SMPL
CCB83 - 1	CCB83	-39	0.0495	-	7470A	PPB	02 Mar 2022 15:38:25	U	SMPL
N1713-04LX5 - 1		69	0.0762	-	7470A	PPB	02 Mar 2022 15:40:41	U	SMPL
N1713-04A - 1		16141	4.0537	-	7470A	PPB	02 Mar 2022 15:42:57	U	SMPL
N1701-01LX5 - 1		-7	0.0574	-	7470A	PPB	02 Mar 2022 15:45:13	U	SMPL
N1701-01A - 1		16854	4.2302	-	7470A	PPB	02 Mar 2022 15:47:28	U	SMPL
CCV84 - 1	CCV84	19108	4.7880	-	7470A	PPB	02 Mar 2022 15:49:44	U	SMPL
CCB84 - 1	CCB84	34	0.0675	-	7470A	PPB	02 Mar 2022 15:52:00	U	SMPL