



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118958

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/02/2022	15:50
2	CAL2	1	Water	NA	NA	20.7	7.00	03/02/2022	15:51
3	CAL3	1	Water	NA	NA	20.7	10.03	03/02/2022	17:55
4	ICV	1	Water	NA	NA	20.7	6.99	03/02/2022	16:00
5	CCV1	1	Water	NA	NA	20.6	2.01	03/02/2022	16:05
6	N1717-02	1	Solid	20.03	20	20.3	9.23	03/02/2022	16:10
7	N1717-02DUP	1	Solid	20.04	20	20.4	9.24	03/02/2022	16:11
8	N1750-02	1	Solid	20.02	20	20.9	8.49	03/02/2022	16:15
9	CCV2	1	Water	NA	NA	20.6	12.03	03/02/2022	16:20

118958

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY N1750 WorkList ID : 157333 Department : Wet-Chemistry Date : 03-02-2022 15:34:52

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
03/01/2022	Solid	N1717-02	Corrosivity	Cool 4 deg C	PSEG03	P11 ETGI-359	03/01/2022	9045D
03/02/2022	Solid	N1750-02	Corrosivity	Cool 4 deg C	DVIR01	P11 WC-2	03/02/2022	9045D

Date/Time 03/02/22 15:43
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: RJ CETA-1060

Date/Time 03/02/22 18:00
Raw Sample Received by: RJ CETA-1060
Raw Sample Relinquished by: JD (WC)