



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118959

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/02/2022	15:50
2	CAL2	1	Water	NA	NA	20.7	7.00	03/02/2022	15:51
3	CAL3	1	Water	NA	NA	20.7	10.03	03/02/2022	17:55
4	ICV	1	Water	NA	NA	20.7	6.99	03/02/2022	16:00
5	CCV1	1	Water	NA	NA	20.8	2.02	03/02/2022	16:20
6	N1748-01	1	Solid	20.02	20	20.2	4.31	03/02/2022	16:22
7	N1748-01DUP	1	Solid	20.01	20	20.3	4.32	03/02/2022	16:23
8	N1748-02	1	Solid	20.03	20	20.8	4.39	03/02/2022	16:25
9	CCV2	1	Water	NA	NA	20.7	12.03	03/02/2022	16:26

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n1748 s

WorkList ID : 157337

Department : Wet-Chemistry

Date : 03-02-2022 16:55:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/02/2022	Solid	N1748-01	pH	Cool 4 deg C	PSEG03	P11	31462	03/02/2022	9045D
03/02/2022	Solid	N1748-02	pH	Cool 4 deg C	PSEG03	P11	31463	03/02/2022	9045D

Date/Time 03/02/22 16:10
 Raw Sample Received by: RD (Ext-1466)
 Raw Sample Relinquished by: RD (Ext-1466)

Date/Time 03/02/22 18:10
 Raw Sample Received by: RD (Ext-1466)
 Raw Sample Relinquished by: RD (Ext-1466)