



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118960

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.02	03/02/2022	16:30
2	CAL2	1	Water	NA	NA	20.8	7.00	03/02/2022	16:31
3	CAL3	1	Water	NA	NA	20.8	10.03	03/02/2022	16:32
4	ICV	1	Water	NA	NA	20.7	6.98	03/02/2022	16:35
5	CCV1	1	Water	NA	NA	20.4	2.01	03/02/2022	17:36
6	N1735-04	1	Water	NA	NA	20.7	5.26	03/02/2022	17:40
7	N1749-01	1	Water	NA	NA	20.4	7.78	03/02/2022	17:42
8	N1752-01	1	Water	NA	NA	20.6	5.06	03/02/2022	17:45
9	N1753-01	1	Water	NA	NA	20.2	6.87	03/02/2022	17:50
10	N1756-02	1	Water	NA	NA	20.1	7.20	03/02/2022	17:52
11	N1756-04	1	Water	NA	NA	20.2	7.24	03/02/2022	18:00
12	N1756-05	1	Water	NA	NA	20.3	6.86	03/02/2022	18:02
13	N1756-06	1	Water	NA	NA	20.3	7.13	03/02/2022	18:05
14	N1757-01	1	Water	NA	NA	20.2	6.03	03/02/2022	18:06
15	N1757-01DUP	1	Water	NA	NA	20.3	6.04	03/02/2022	18:07
16	CCV2	1	Water	NA	NA	20.6	12.03	03/02/2022	18:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1752

WorkList ID : 157328

Department : Wet-Chemistry

Date : 03-02-2022 15:02:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/02/2022	Water	N1735-04	pH	Cool 4 deg C	PSEG03	P11	34061	03/02/2022	9040C
03/02/2022	Water	N1749-01	pH	Cool 4 deg C	PSEG04	P11	402	03/02/2022	9040C
02/25/2022	Water	N1752-01	pH	Cool 4 deg C	STER04	P11	POOL-WATER	02/25/2022	9040C
03/02/2022	Water	N1753-01	pH	Cool 4 deg C	PSEG03	P11	31465	03/02/2022	9040C
03/02/2022	Water	N1756-02	pH	Cool 4 deg C	PSEG03	P11	33579	03/02/2022	9040C
03/02/2022	Water	N1756-04	pH	Cool 4 deg C	PSEG03	P11	33569	03/02/2022	9040C
03/02/2022	Water	N1756-05	pH	Cool 4 deg C	PSEG03	P11	33568	03/02/2022	9040C
03/02/2022	Water	N1756-06	pH	Cool 4 deg C	PSEG03	P11	33567	03/02/2022	9040C
03/02/2022	Water	N1757-01	pH	Cool 4 deg C	PSEG03	P11	34064	03/02/2022	9040C

Date/Time 03/02/22 16:10
 Raw Sample Received by: JO LWC
 Raw Sample Relinquished by: RJ C E L L C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

JO 18:15
 03/02/22 18:00 JO
 RJ C E L L C
 JO LWC

VP 118960