


LB118965

7471B

INSTRUMENT ID: CV8

 μ Abs.:

52079



Linear

A= 0.0000e+000

B= 1.9082e-004

C= -9.3123e-003

Rho= 0.9998628

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Good
0.0	0.000	-0.009	-0.009	2	0.000	2					1
0.2	0.200	0.167	-0.033	925	0.0 %	925					-17
2.5	2.500	2.469	-0.031	12990	0.0 %	12990					-1
5.0	5.000	5.121	0.121	26886	0.0 %	26886					2
7.5	7.500	7.523	0.023	39474	0.0 %	39474					0
10.0	10.000	9.928	-0.072	52079	0.0 %	52079					-1

LB118965 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 03 Mar 2022 09:15:17							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1	50	2	-	0.0000	7471B	PPB	03 Mar 2022 09:17:39	S	Std		
0.2 - 1	50.2	925	-	0.2000	7471B	PPB	03 Mar 2022 09:19:57	S	Std		
2.5 - 1	52.5	12990	-	2.5000	7471B	PPB	03 Mar 2022 09:22:14	S	Std		
5.0 - 1	55	26886	-	5.0000	7471B	PPB	03 Mar 2022 09:24:31	S	Std		
7.5 - 1	57.5	39474	-	7.5000	7471B	PPB	03 Mar 2022 09:26:48	S	Std		
10.0 - 1	510	52079	-	10.0000	7471B	PPB	03 Mar 2022 09:31:39	S	Std		
ICV16 - 1	ICV16	20527	3.9076	-	7471B	PPB	03 Mar 2022 09:34:24	U	SMPL		
ICB16 - 1	ICB16	-60	-0.0208	-	7471B	PPB	03 Mar 2022 09:36:40	U	SMPL		
CCV85 - 1	CCV85	26965	5.1361	-	7471B	PPB	03 Mar 2022 09:38:55	U	SMPL		
CCB85 - 1	CCB85	-57	-0.0202	-	7471B	PPB	03 Mar 2022 09:41:11	U	SMPL		
CRA - 1	CRA	916	0.1655	-	7471B	PPB	03 Mar 2022 09:43:26	U	SMPL		
HighStd - 1	HighStd	52671	10.0412	-	7471B	PPB	03 Mar 2022 09:45:41	U	SMPL		
ChkStd - 1	ChkStd	36526	6.9605	-	7471B	PPB	03 Mar 2022 09:47:57	U	SMPL		
PB143048BL - 1	PBS	-91	-0.0267	-	7471B	PPB	03 Mar 2022 09:50:15	U	SMPL		
PB143048BS - 1	LCSS	20693	3.9393	-	7471B	PPB	03 Mar 2022 09:52:32	U	SMPL		
N1717-01 - 1	ETGI-359	3689	0.6946	-	7471B	PPB	03 Mar 2022 09:54:49	U	SMPL		
N1719-01 - 1	TT-148-IDWSO-20220228	192	0.0273	-	7471B	PPB	03 Mar 2022 09:57:05	U	SMPL		
N1721-01 - 1	DUM-1	2354	0.4399	-	7471B	PPB	03 Mar 2022 09:59:23	U	SMPL		
N1721-03 - 1	DUM-2	380	0.0632	-	7471B	PPB	03 Mar 2022 10:01:40	U	SMPL		
N1721-05 - 1	DUM-3	1103	0.2012	-	7471B	PPB	03 Mar 2022 10:03:55	U	SMPL		
CCV86 - 1	CCV86	27052	5.1527	-	7471B	PPB	03 Mar 2022 10:06:11	U	SMPL		
CCB86 - 1	CCB86	-55	-0.0198	-	7471B	PPB	03 Mar 2022 10:08:26	U	SMPL		
N1721-07 - 1	DUM-4	1849	0.3435	-	7471B	PPB	03 Mar 2022 10:10:42	U	SMPL		
N1721-09 - 1	DUM-5	365	0.0603	-	7471B	PPB	03 Mar 2022 10:12:57	U	SMPL		
N1722-01 - 1	DUM-6	1526	0.2819	-	7471B	PPB	03 Mar 2022 10:15:13	U	SMPL		
N1722-03 - 1	DUM-7	7050	1.3360	-	7471B	PPB	03 Mar 2022 10:17:31	U	SMPL		
N1722-05 - 1	DUM-8	633	0.1115	-	7471B	PPB	03 Mar 2022 10:19:47	U	SMPL		
N1722-07 - 1	DUM-9	6015	1.1385	-	7471B	PPB	03 Mar 2022 10:22:03	U	SMPL		
N1722-09 - 1	DUM-10	1883	0.3500	-	7471B	PPB	03 Mar 2022 10:24:20	U	SMPL		
N1722-09DUP - 1	DUM-10DUP	1837	0.3412	-	7471B	PPB	03 Mar 2022 10:31:11	U	SMPL		
N1722-09MS - 1	DUM-10MS	22025	4.1934	-	7471B	PPB	03 Mar 2022 10:33:28	U	SMPL		
N1722-09MSD - 1	DUM-10MSD	22034	4.1952	-	7471B	PPB	03 Mar 2022 10:35:44	U	SMPL		
CCV87 - 1	CCV87	26825	5.1094	-	7471B	PPB	03 Mar 2022 10:37:59	U	SMPL		
CCB87 - 1	CCB87	-42	-0.0173	-	7471B	PPB	03 Mar 2022 10:40:15	U	SMPL		
N1722-09LX5 - 1		401	0.0672	-	7471B	PPB	03 Mar 2022 10:42:31	U	SMPL		
N1722-09A - 1		23921	4.5552	-	7471B	PPB	03 Mar 2022 10:44:47	U	SMPL		
CCV88 - 1	CCV88	26534	5.0538	-	7471B	PPB	03 Mar 2022 10:47:03	U	SMPL		
CCB88 - 1	CCB88	-42	-0.0173	-	7471B	PPB	03 Mar 2022 10:49:18	U	SMPL		