



Analytical Summary Report

Analysis Method: SM2120 B

Reviewed By: Iwona

Parameter: Color

Supervisor Review By: apatel

Run Number: LB118969

pH Meter ID: WC pH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP95670
Calibration std, 5 color unit	WP95669
Calibration std, 10 color unit	WP95668
Calibration std, 15 color unit	WP95667
Calibration std, 20 color unit	WP95666
Calibration std, 25 color unit	WP95665
Calibration std, 30 color unit	WP95664
Calibration std, 40 color unit	WP95663
Calibration std, 50 color unit	WP95662
Calibration std, 60 color unit	WP95661
Calibration std, 70 color unit	WP95660

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB118969BL	LB118969BL	1	5.12	N/A	0.00	03/03/2022	10:30
2	N1746-01	MONTHLY-OUTFALL-002	10	8.35	N/A	70.00	03/03/2022	10:35
3	N1746-01DUP	MONTHLY-OUTFALL-002	10	8.32	N/A	70.00	03/03/2022	10:40

WORKLIST(Hardcopy Internal Chain)

L131184 69

WorkList Name : COLOR-030322 WorkList ID : 157354 Department : Wet-Chemistry Date : 03-03-2022 09:16:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/04/2022	Water	N1746-01	Color	Cool 4 deg C	INFR01	P11	MONTHLY-OUTFALL-002	03/02/2022	SM2120 B

Date/Time 3/3/22 10:20
 Raw Sample Received by: JD (WCT)
 Raw Sample Relinquished by: JD (WCT)

Date/Time 3/3/22
 Raw Sample Received by: JD (WCT)
 Raw Sample Relinquished by: JD (WCT)