



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118972

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/03/2022	10:45
2	CAL2	1	Water	NA	NA	20.3	7.00	03/03/2022	10:46
3	CAL3	1	Water	NA	NA	20.3	10.03	03/03/2022	10:47
4	ICV	1	Water	NA	NA	20.4	7.00	03/03/2022	10:48
5	CCV1	1	Water	NA	NA	20.2	2.02	03/03/2022	10:49
6	N1742-01	1	Water	NA	NA	20.1	7.44	03/03/2022	10:52
7	N1769-02	1	Water	NA	NA	20.2	9.51	03/03/2022	11:00
8	N1769-03	1	Water	NA	NA	20.3	10.51	03/03/2022	11:05
9	N1769-03DUP	1	Water	NA	NA	20.4	10.52	03/03/2022	11:06
10	CCV2	1	Water	NA	NA	20.6	12.03	03/03/2022	11:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1769 WorkList ID : 157369 Department : Wet-Chemistry Date : 03-03-2022 10:39:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/02/2022	Water	N1742-01	pH	Cool 4 deg C	INFR01	P11	BI-WEEKLY-INFLUENT	03/02/2022	9040C
03/02/2022	Water	N1769-02	pH	Cool 4 deg C	TETR06	P11	TT-036-IDWGW-20220302	03/02/2022	9040C
03/02/2022	Water	N1769-03	pH	Cool 4 deg C	TETR06	P11	TT-037-IDWGW-20220302	03/02/2022	9040C

Date/Time 03/03/22 10:40
 Raw Sample Received by: JP (WCI)
 Raw Sample Relinquished by: JP (SW)

Date/Time 03/03/22 12:30
 Raw Sample Received by: JP (SW)
 Raw Sample Relinquished by: JP