



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118990

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	03/03/2022	09:50
2	N1748-03		1	10	100.00	>212.0	>212.0	03/03/2022	10:15
3	N1748-03DUP		1	10	100.00	>212.0	>212.0	03/03/2022	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB118990

WorkList Name : fp3322 WorkList ID : 157358 Department : Wet-Chemistry Date : 03-03-2022 09:28:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/15/2022	Solid	N1748-03	Flash Point	Cool 4 deg C	PSEG03	P11	31462-63	03/02/2022	1010A

Date/Time 03/03/22 09:30
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: NB(wc)

Date/Time 03/03/22 10:50
 Raw Sample Received by: NB(wc)
 Raw Sample Relinquished by: KS(wc)