



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118991

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/03/2022	16:45
2	CAL2	1	Water	NA	NA	20.4	7.01	03/03/2022	16:46
3	CAL3	1	Water	NA	NA	20.4	10.03	03/03/2022	16:47
4	ICV	1	Water	NA	NA	20.3	6.99	03/03/2022	16:50
5	CCV1	1	Water	NA	NA	20.5	2.01	03/03/2022	16:52
6	N1782-02	1	Water	NA	NA	20.4	7.13	03/03/2022	17:00
7	N1782-02DUP	1	Water	NA	NA	20.5	7.14	03/03/2022	17:01
8	N1782-04	1	Water	NA	NA	20.6	7.08	03/03/2022	17:03
9	N1782-06	1	Water	NA	NA	20.6	6.93	03/03/2022	17:05
10	N1787-01	1	Water	NA	NA	20.4	6.22	03/03/2022	17:09
11	N1789-03	1	Water	NA	NA	20.2	6.46	03/03/2022	17:10
12	CCV2	1	Water	NA	NA	20.5	12.03	03/03/2022	17:11

WORKLIST(Hardcopy Internal Chain)

13118991

WorkList Name : PH N1746

WorkList ID : 157388

Department : Wet-Chemistry

Date : 03-03-2022 14:43:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/03/2022	Water	N1782-02	pH A	Cool 4 deg C	PSEG04	P11	DSN002	03/03/2022	9040C
03/03/2022	Water	N1782-04	pH A	Cool 4 deg C	PSEG04	P11	DSN001	03/03/2022	9040C
03/03/2022	Water	N1782-06	pH A	Cool 4 deg C	PSEG04	P11	RWS003	03/03/2022	9040C
03/03/2022	Water	N1787-01	pH A	Cool 4 deg C	PSEG03	P11	WATER-DRUM-1-3	03/03/2022	9040C
03/03/2022	Water	N1789-03	pH A	Cool 4 deg C	PSEG03	P11	786	03/03/2022	9040C

Date/Time 03/03/22 16:35

Raw Sample Received by: JO (WC)

Raw Sample Relinquished by: KS(WC)

Date/Time 03/03/22

Raw Sample Received by: KS(WC)

Raw Sample Relinquished by: JO(WC)