


LB 118994

7471B

INSTRUMENT ID: CV 2

 μ Abs.:

46594



Linear

A= 0.0000e+000

B= 2.1348e-004

C= -1.8961e-002

Rho= 0.9997222

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.018	-0.018	4	0.000	4					-
0.2	0.200	0.200	-0.000	1024	0.0 %	1024					0
2.5	2.500	2.411	-0.089	11384	0.0 %	11384					-4
5.0	5.000	5.179	0.179	24348	0.0 %	24348					4
7.5	7.500	7.500	0.000	35223	0.0 %	35223					0
10.0	10.000	9.928	-0.072	46594	0.0 %	46594					-1

LB118994 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Mar 2022 09:23:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		4	-	0.0000	7471B	PPB	04 Mar 2022 09:30:41	S	Std
0.2 - 1		1024	-	0.2000	7471B	PPB	04 Mar 2022 09:32:58	S	Std
2.5 - 1		11384	-	2.5000	7471B	PPB	04 Mar 2022 09:35:15	S	Std
5.0 - 1		24348	-	5.0000	7471B	PPB	04 Mar 2022 09:37:32	S	Std
7.5 - 1		35223	-	7.5000	7471B	PPB	04 Mar 2022 09:39:49	S	Std
10.0 - 1		46594	-	10.0000	7471B	PPB	04 Mar 2022 09:42:07	S	Std
ICV19 - 1	ICV19	18270	3.8813	-	7471B	PPB	04 Mar 2022 09:44:53	U	SMPL
ICB19 - 1	ICB19	-59	-0.0316	-	7471B	PPB	04 Mar 2022 09:47:08	U	SMPL
CCV01 - 1	CCV01	23388	4.9739	-	7471B	PPB	04 Mar 2022 09:49:25	U	SMPL
CCB01 - 1	CCB01	-60	-0.0318	-	7471B	PPB	04 Mar 2022 09:51:40	U	SMPL
CRA - 1	CRA	917	0.1768	-	7471B	PPB	04 Mar 2022 09:53:57	U	SMPL
HighStd - 1	HighStd	45523	9.6993	-	7471B	PPB	04 Mar 2022 09:56:14	U	SMPL
ChkStd - 1	ChkStd	31319	6.6670	-	7471B	PPB	04 Mar 2022 09:58:29	U	SMPL
PB143087BL - 1	PBW	-99	-0.0401	-	7471B	PPB	04 Mar 2022 10:00:47	U	SMPL
PB143087BS - 1	LCSW	17047	3.6202	-	7471B	PPB	04 Mar 2022 10:03:05	U	SMPL
N1731-03 - 1	VACLOT-TOPSOIL-GRILLO-01	16	-0.0155	-	7471B	PPB	04 Mar 2022 10:05:22	U	SMPL
N1731-04 - 1	VACLOT-TOPSOIL-GRILLO-01A	34	-0.0117	-	7471B	PPB	04 Mar 2022 10:07:39	U	SMPL
N1731-04DUP - 1	VACLOT-TOPSOIL-GRILLO-01ADUP	23	-0.0141	-	7471B	PPB	04 Mar 2022 10:09:56	U	SMPL
N1731-04MS - 1	VACLOT-TOPSOIL-GRILLO-01AMS	17034	3.6174	-	7471B	PPB	04 Mar 2022 10:12:14	U	SMPL
N1731-04MSD - 1	VACLOT-TOPSOIL-GRILLO-01AMSD	17048	3.6204	-	7471B	PPB	04 Mar 2022 10:14:29	U	SMPL
CCV02 - 1	CCV02	23775	5.0565	-	7471B	PPB	04 Mar 2022 10:16:46	U	SMPL
CCB02 - 1	CCB02	-73	-0.0345	-	7471B	PPB	04 Mar 2022 10:19:02	U	SMPL
PB143088BL - 1	PBW	-1	-0.0192	-	7471B	PPB	04 Mar 2022 10:21:19	U	SMPL
PB143088BS - 1	LCSW	16924	3.5940	-	7471B	PPB	04 Mar 2022 10:23:35	U	SMPL
N1734-02 - 1	COMP	167	0.0167	-	7471B	PPB	04 Mar 2022 10:25:50	U	SMPL
N1769-02 - 1	TT-036-IDWGW-20220302	117	0.0060	-	7471B	PPB	04 Mar 2022 10:28:07	U	SMPL
N1769-02DUP - 1	TT-036-IDWGW-20220302DUP	126	0.0079	-	7471B	PPB	04 Mar 2022 10:30:23	U	SMPL
N1769-02MS - 1	TT-036-IDWGW-20220302MS	17832	3.7878	-	7471B	PPB	04 Mar 2022 10:32:40	U	SMPL
N1769-02MSD - 1	TT-036-IDWGW-20220302MSD	18339	3.8960	-	7471B	PPB	04 Mar 2022 10:34:57	U	SMPL
N1769-03 - 1	TT-037-IDWGW-20220302	-43	-0.0281	-	7471B	PPB	04 Mar 2022 10:37:14	U	SMPL
PB143087TB - 1	PB143046TB	21	-0.0145	-	7471B	PPB	04 Mar 2022 10:39:31	U	SMPL
CCV03 - 1	CCV03	23603	5.0198	-	7471B	PPB	04 Mar 2022 10:41:47	U	SMPL
CCB03 - 1	CCB03	-70	-0.0339	-	7471B	PPB	04 Mar 2022 10:44:03	U	SMPL
N1731-04LX5 - 1		84	-0.0010	-	7471B	PPB	04 Mar 2022 10:46:20	U	SMPL
N1731-04A - 1		18564	3.9441	-	7471B	PPB	04 Mar 2022 10:48:36	U	SMPL
N1769-02LX5 - 1		-29	-0.0252	-	7471B	PPB	04 Mar 2022 10:50:52	U	SMPL
N1769-02A - 1		19483	4.1403	-	7471B	PPB	04 Mar 2022 10:53:09	U	SMPL
CCV04 - 1	CCV04	23555	5.0095	-	7471B	PPB	04 Mar 2022 10:55:24	U	SMPL
CCB04 - 1	CCB04	-83	-0.0367	-	7471B	PPB	04 Mar 2022 10:57:41	U	SMPL