



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118997

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	03/04/2022	09:35
2	N1712-03		1	12	100.00	>212.0	>212.0	03/04/2022	09:58
3	N1712-03DUP		1	12	100.00	>212.0	>212.0	03/04/2022	10:20
4	N1735-04		1	15	100.00	>212.0	>212.0	03/04/2022	10:45
5	N1753-01		1	13	100.00	>212.0	>212.0	03/04/2022	11:10
6	N1756-01		1	14	100.00	>212.0	>212.0	03/04/2022	11:40
7	N1756-03		1	15	100.00	>212.0	>212.0	03/04/2022	12:05
8	N1757-01		1	12	100.00	>212.0	>212.0	03/04/2022	12:32
9	N1787-01		1	14	100.00	>212.0	>212.0	03/04/2022	12:55
10	N1789-03		1	14	100.00	>212.0	>212.0	03/04/2022	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB118997

WorkList Name : fp3222 WorkList ID : 157315 Department : Wet-Chemistry Date : 03-02-2022 10:44:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/07/2022	Water	N1712-03	C Flash Point	Cool 4 deg C	PSEG03	P11	31404	02/28/2022	1010B
03/09/2022	Water	N1735-04	✓ Flash Point	Cool 4 deg C	PSEG03	P11	34061	03/02/2022	1010B
03/15/2022	Water	N1753-01	K Flash Point	Cool 4 deg C	PSEG03	P11	31465	03/02/2022	1010B
03/09/2022	Water	N1756-01	D Flash Point	Cool 4 deg C	PSEG03	P11	33568-33579	03/02/2022	1010B
03/09/2022	Water	N1756-03	C Flash Point	Cool 4 deg C	PSEG03	P11	33567-33569	03/02/2022	1010B
03/09/2022	Water	N1757-01	K Flash Point	Cool 4 deg C	PSEG03	P11	34064	03/02/2022	1010B
03/10/2022	Water	N1787-01	C Flash Point	Cool 4 deg C	PSEG03	P11	WATER-DRUM-1-3	03/03/2022	1010B
03/10/2022	Water	N1789-03	C Flash Point	Cool 4 deg C	PSEG03	P11	786	03/03/2022	1010B

Date/Time 03/04/22 09:15
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: RJ CFA-1ab

Date/Time 03/04/22 13:20
 Raw Sample Received by: RJ CFA-1ab
 Raw Sample Relinquished by: KS(WC)