



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119001

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/04/2022	11:00
2	CAL2	1	Water	NA	NA	20.3	7.01	03/04/2022	11:01
3	CAL3	1	Water	NA	NA	20.3	10.03	03/04/2022	11:02
4	ICV	1	Water	NA	NA	20.4	6.98	03/04/2022	11:05
5	CCV1	1	Water	NA	NA	20.6	2.01	03/04/2022	11:06
6	N1799-01	1	Water	NA	NA	20.4	8.90	03/04/2022	11:10
7	N1799-01DUP	1	Water	NA	NA	20.5	8.91	03/04/2022	11:11
8	CCV2	1	Water	NA	NA	20.4	12.03	03/04/2022	11:15

2B119001

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1799

WorkList ID : 157421

Department : Wet-Chemistry

Date : 03-04-2022 10:40:38

Due Date		Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/03/2022	Water	N1799-01	pH	D	Cool 4 deg C	TETR06	N31	TT-038-1DWGW-20220303	03/03/2022	9040C

Date/Time 03/04/22 10:50
Raw Sample Received by: JD (CPC)
Raw Sample Relinquished by: JD (CPC)

Date/Time 03/04/22 12:00
Raw Sample Received by: JD (CPC)
Raw Sample Relinquished by: JD (CPC)