


LB119006

7471B

INSTRUMENT ID: CV1

 μ Abs.:

5388E



Linear

A= 0.0000e+000

B= 1.8451e-004

C= -1.3934e-002

Rho= 0.9999270

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	-0.011	-0.011	15	0.000	15					
0.2	0.200	0.175	-0.025	1023	0.0 %	1023					-13
2.5	2.500	2.503	0.003	13641	0.0 %	13641					0
5.0	5.000	5.035	0.035	27363	0.0 %	27363					1
7.5	7.500	7.570	0.070	41103	0.0 %	41103					1
10.0	10.000	9.929	-0.071	53888	0.0 %	53888					-1

LB119006 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Mar 2022 12:32:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.0000	7471B	PPB	04 Mar 2022 12:34:33	S	Std
0.2 - 1	50.2	1023	-	0.2000	7471B	PPB	04 Mar 2022 12:36:51	S	Std
2.5 - 1	52.5	13641	-	2.5000	7471B	PPB	04 Mar 2022 12:39:08	S	Std
5.0 - 1	55	27363	-	5.0000	7471B	PPB	04 Mar 2022 12:41:25	S	Std
7.5 - 1	57.5	41103	-	7.5000	7471B	PPB	04 Mar 2022 12:54:01	S	Std
10.0 - 1	510	53888	-	10.0000	7471B	PPB	04 Mar 2022 12:56:17	S	Std
ICV20 - 1	ICV20	19931	3.6635	-	7471B	PPB	04 Mar 2022 12:58:59	U	SMPL
ICB20 - 1	ICB20	-66	-0.0261	-	7471B	PPB	04 Mar 2022 13:01:14	U	SMPL
CCV05 - 1	CCV05	27040	4.9751	-	7471B	PPB	04 Mar 2022 13:08:44	U	SMPL
CCB05 - 1	CCB05	-64	-0.0257	-	7471B	PPB	04 Mar 2022 13:10:59	U	SMPL
CRA - 1	CRA	1023	0.1748	-	7471B	PPB	04 Mar 2022 13:13:15	U	SMPL
HighStd - 1	HighStd	53045	9.7733	-	7471B	PPB	04 Mar 2022 13:15:30	U	SMPL
ChkStd - 1	ChkStd	40116	7.3878	-	7471B	PPB	04 Mar 2022 13:17:45	U	SMPL
PB143118BL - 1	PBS	-68	-0.0265	-	7471B	PPB	04 Mar 2022 13:20:03	U	SMPL
PB143118BS - 1	LCSS	23495	4.3211	-	7471B	PPB	04 Mar 2022 13:22:20	U	SMPL
N1762-01 - 1	282	1126	0.1938	-	7471B	PPB	04 Mar 2022 13:24:36	U	SMPL
N1774-01 - 1	DRUM-COMP-A	23336	4.2917	-	7471B	PPB	04 Mar 2022 13:26:53	U	SMPL
N1774-03 - 1	DRUM-COMP-B	2845	0.5110	-	7471B	PPB	04 Mar 2022 13:29:10	U	SMPL
N1785-01 - 1	TR-4-030322	890	0.1503	-	7471B	PPB	04 Mar 2022 13:31:28	U	SMPL
N1789-01 - 1	CHRT-27165	651	0.1062	-	7471B	PPB	04 Mar 2022 13:33:43	U	SMPL
CCV06 - 1	CCV06	27083	4.9831	-	7471B	PPB	04 Mar 2022 13:35:59	U	SMPL
CCB06 - 1	CCB06	-54	-0.0239	-	7471B	PPB	04 Mar 2022 13:38:14	U	SMPL
N1790-01 - 1	DUM-17	2303	0.4110	-	7471B	PPB	04 Mar 2022 13:40:30	U	SMPL
N1790-03 - 1	DUM-18	2717	0.4874	-	7471B	PPB	04 Mar 2022 13:42:46	U	SMPL
N1790-05 - 1	DUM-19	1402	0.2447	-	7471B	PPB	04 Mar 2022 13:45:02	U	SMPL
N1790-07 - 1	DUM-20	1967	0.3490	-	7471B	PPB	04 Mar 2022 13:47:17	U	SMPL
N1790-09 - 1	DUM-21	802	0.1340	-	7471B	PPB	04 Mar 2022 13:49:33	U	SMPL
N1790-09DUP - 1	DUM-21DUP	711	0.1173	-	7471B	PPB	04 Mar 2022 13:51:50	U	SMPL
N1790-09MS - 1	DUM-21MS	22149	4.0727	-	7471B	PPB	04 Mar 2022 13:54:06	U	SMPL
N1790-09MSD - 1	DUM-21MSD	21831	4.0140	-	7471B	PPB	04 Mar 2022 13:56:23	U	SMPL
N1790-09LX5 - 1		102	0.0049	-	7471B	PPB	04 Mar 2022 13:58:41	U	SMPL
N1790-09A - 1		23507	4.3233	-	7471B	PPB	04 Mar 2022 14:00:57	U	SMPL
CCV07 - 1	CCV07	27239	5.0119	-	7471B	PPB	04 Mar 2022 14:03:13	U	SMPL
CCB07 - 1	CCB07	-54	-0.0239	-	7471B	PPB	04 Mar 2022 14:05:29	U	SMPL