

LB119012

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

3/4/2022 13:37

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.755	0.0	0.082	
ICB1	-0.004	0.0	0.001	
CCV1	249.050	0.0	0.213	
CCB1	-0.010	0.0	0.001	
PB143089BL	-0.774	0.0	0.001	
N1746-02	1.503	0.0	0.003	
N1746-02DUP	1.366	0.0	0.002	
CCV2	258.326	0.0	0.220	
CCB2	-0.151	0.0	0.001	

N	9
Mean	67.118
SD	110.2260
CV%	164.23

Aquakem v. 7.2AQ1

Results from time period:

Fri Mar 04 12:23:44 2022

Fri Mar 04 13:30:17 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.0718	µg/l	3/4/2022 12:58:02	
5.0PPBCN	A	Total CN	P	4.2172	µg/l	3/4/2022 12:58:03	
10PPBCN	A	Total CN	P	9.6776	µg/l	3/4/2022 12:58:04	
50PPBCN	A	Total CN	P	48.535	µg/l	3/4/2022 12:58:05	
100PPBCN	A	Total CN	P	99.285	µg/l	3/4/2022 12:58:06	
250PPBCN	A	Total CN	P	256.1065	µg/l	3/4/2022 12:58:07	
500PPBCN	A	Total CN	P	497.2505	µg/l	3/4/2022 12:58:08	
ICV1	S	Total CN	P	94.7547	µg/l	3/4/2022 13:26:25	
ICB1	S	Total CN	P	-0.0042	µg/l	3/4/2022 13:26:27	
CCV1	S	Total CN	P	249.0503	µg/l	3/4/2022 13:26:29	
CCB1	S	Total CN	P	-0.0097	µg/l	3/4/2022 13:26:30	
PB143089BL	S	Total CN	P	-0.7743	µg/l	3/4/2022 13:26:33	
N1746-02	S	Total CN	P	1.5029	µg/l	3/4/2022 13:26:34	
N1746-02DUP	S	Total CN	P	1.3662	µg/l	3/4/2022 13:30:15	
CCV2	S	Total CN	P	258.326	µg/l	3/4/2022 13:30:16	
CCB2	S	Total CN	P	-0.1506	µg/l	3/4/2022 13:30:17	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

3/4/2022 13:04


Test Total CN

Accepted 3/4/2022 13:04

Factor 1179
Bias 0.001

Coeff. of det. 0.999765

Errors



	Calibrator	Response	Calc. con.	Conc.	^{Re} Errors
1	0.0PPBCN	0.001	-0.0718	0.0000	-
2	5.0PPBCN	0.005	4.2172	5.0000	-15.7
3	10PPBCN	0.009	9.6776	10.0000	-3.2
4	50PPBCN	0.042	48.5350	50.0000	-2.9
5	100PPBCN	0.086	99.2850	100.0000	-0.7
6	250PPBCN	0.219	256.1065	250.0000	2.4
7	500PPBCN	0.423	497.2505	500.0000	-0.5

12 3/4/22