

LB119014


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34423



A= 0.0000e+000

B= 2.9554e-004

C= 1.9427e-002

Rho= 0.9993210

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/01
0.0	0.000	0.032	0.032	44	0.000	44					16
0.2	0.200	0.232	0.032	718	0.0 %	718					2
2.5	2.500	2.542	0.042	8537	0.0 %	8537					-1
5.0	5.000	4.963	-0.037	16726	0.0 %	16726					-3
7.5	7.500	7.238	-0.262	24426	0.0 %	24426					2
10.0	10.000	10.193	0.193	34423	0.0 %	34423					

LB119014 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Mar 2022 14:54:05

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	44	-	0.0000	7470A	PPB	04 Mar 2022 14:54:45	S	Std
0.2 - 1	50-2	718	-	0.2000	7470A	PPB	04 Mar 2022 14:57:03	S	Std
2.5 - 1	52-1	8537	-	2.5000	7470A	PPB	04 Mar 2022 14:59:20	S	Std
5.0 - 1	55	16726	-	5.0000	7470A	PPB	04 Mar 2022 15:01:37	S	Std
7.5 - 1	59-1	24426	-	7.5000	7470A	PPB	04 Mar 2022 15:06:34	S	Std
10.0 - 1	510	34423	-	10.0000	7470A	PPB	04 Mar 2022 15:08:50	S	Std
ICV21 - 1	ICV21	14039	4.1685	-	7470A	PPB	04 Mar 2022 15:11:35	U	SMPL
ICB21 - 1	ICB21	-28	0.0112	-	7470A	PPB	04 Mar 2022 15:13:50	U	SMPL
CCV08 - 1	CCV08	17603	5.2218	-	7470A	PPB	04 Mar 2022 15:16:05	U	SMPL
CCB08 - 1	CCB08	-29	0.0109	-	7470A	PPB	04 Mar 2022 15:18:21	U	SMPL
CRA - 1	CRA	679	0.2201	-	7470A	PPB	04 Mar 2022 15:20:36	U	SMPL
HighStd - 1	HighStd	34027	10.0757	-	7470A	PPB	04 Mar 2022 15:22:51	U	SMPL
ChkStd - 1	ChkStd	23171	6.8673	-	7470A	PPB	04 Mar 2022 15:25:06	U	SMPL
PB143109BL - 1	PBW	-50	0.0047	-	7470A	PPB	04 Mar 2022 15:27:23	U	SMPL
PB143109BS - 1	LCSW	13800	4.0978	-	7470A	PPB	04 Mar 2022 15:32:09	U	SMPL
N1732-01 - 1	BLDG39-CONSTRUCTION-DEBRIS	-21	0.0132	-	7470A	PPB	04 Mar 2022 15:34:25	U	SMPL
N1732-02 - 1	BLDG39-PLASTER-DEBRIS	53	0.0351	-	7470A	PPB	04 Mar 2022 15:36:42	U	SMPL
N1748-03 - 1	31462-63	196	0.0774	-	7470A	PPB	04 Mar 2022 15:38:59	U	SMPL
N1750-02 - 1	WC-2	54	0.0354	-	7470A	PPB	04 Mar 2022 15:41:16	U	SMPL
N1750-02DUP - 1	WC-2DUP	86	0.0448	-	7470A	PPB	04 Mar 2022 15:43:31	U	SMPL
CCV09 - 1	CCV09	17221	5.1089	-	7470A	PPB	04 Mar 2022 15:45:47	U	SMPL
CCB09 - 1	CCB09	-32	0.0100	-	7470A	PPB	04 Mar 2022 15:48:02	U	SMPL
N1750-02MS - 1	WC-2MS	14426	4.2828	-	7470A	PPB	04 Mar 2022 15:52:49	U	SMPL
N1750-02MSD - 1	WC-2MSD	14592	4.3319	-	7470A	PPB	04 Mar 2022 15:55:04	U	SMPL
PB143110BL - 1	PBW	-41	0.0073	-	7470A	PPB	04 Mar 2022 15:57:20	U	SMPL
PB143110BS - 1	LCSW	14483	4.2997	-	7470A	PPB	04 Mar 2022 15:59:36	U	SMPL
N1753-01 - 1	31465	-1	0.0191	-	7470A	PPB	04 Mar 2022 16:01:52	U	SMPL
N1753-01DUP - 1	31465DUP	24	0.0265	-	7470A	PPB	04 Mar 2022 16:04:09	U	SMPL
N1753-01MS - 1	31465MS	14843	4.4061	-	7470A	PPB	04 Mar 2022 16:09:00	U	SMPL
N1753-01MSD - 1	31465MSD	15153	4.4977	-	7470A	PPB	04 Mar 2022 16:11:17	U	SMPL
PB143111BL - 1	PBW	-36	0.0088	-	7470A	PPB	04 Mar 2022 16:13:34	U	SMPL
PB143111BS - 1	LCSW	14624	4.3414	-	7470A	PPB	04 Mar 2022 16:15:50	U	SMPL
CCV10 - 1	CCV10	17327	5.1402	-	7470A	PPB	04 Mar 2022 16:18:06	U	SMPL
CCB10 - 1	CCB10	-36	0.0088	-	7470A	PPB	04 Mar 2022 16:20:22	U	SMPL
N1787-01 - 1	WATER-DRUM-1-3	1026	0.3226	-	7470A	PPB	04 Mar 2022 16:29:34	U	SMPL
N1787-01DUP - 1	WATER-DRUM-1-3DUP	1029	0.3235	-	7470A	PPB	04 Mar 2022 16:31:50	U	SMPL
N1787-01MS - 1	WATER-DRUM-1-3MS	13872	4.1191	-	7470A	PPB	04 Mar 2022 16:34:06	U	SMPL
N1787-01MSD - 1	WATER-DRUM-1-3MSD	12973	3.8534	-	7470A	PPB	04 Mar 2022 16:38:58	U	SMPL
N1789-03 - 1	786	-26	0.0117	-	7470A	PPB	04 Mar 2022 16:41:14	U	SMPL
N1799-01 - 1	TT-038-1DWGW-20220303	310	0.1110	-	7470A	PPB	04 Mar 2022 16:43:31	U	SMPL
PB143077TB - 1	PB143077TB	2	0.0200	-	7470A	PPB	04 Mar 2022 16:45:47	U	SMPL
PB143090TB - 1	PB143090TB	-4	0.0182	-	7470A	PPB	04 Mar 2022 16:48:05	U	SMPL
CCV11 - 1	CCV11	16213	4.8110	-	7470A	PPB	04 Mar 2022 16:50:22	U	SMPL
CCB11 - 1	CCB11	-34	0.0094	-	7470A	PPB	04 Mar 2022 16:52:38	U	SMPL
N1750-02LX5 - 1		60	0.0372	-	7470A	PPB	04 Mar 2022 16:54:54	U	SMPL
N1750-02A - 1		14580	4.3284	-	7470A	PPB	04 Mar 2022 16:57:09	U	SMPL
N1753-01LX5 - 1		-30	0.0106	-	7470A	PPB	04 Mar 2022 16:59:26	U	SMPL
N1753-01A - 1		14384	4.2704	-	7470A	PPB	04 Mar 2022 17:01:41	U	SMPL
N1787-01LX5 - 1		155	0.0652	-	7470A	PPB	04 Mar 2022 17:03:57	U	SMPL
N1787-01A - 1		15169	4.5024	-	7470A	PPB	04 Mar 2022 17:06:13	U	SMPL
CCV12 - 1	CCV12	16554	4.9118	-	7470A	PPB	04 Mar 2022 17:08:30	U	SMPL
CCB12 - 1	CCB12	-43	0.0067	-	7470A	PPB	04 Mar 2022 17:10:46	U	SMPL