



# PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 3/8/2022

OVENTEMP IN Celsius(°C): 103

Time IN: 14:10

In Date: 03/07/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:10

Out Date: 03/08/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB119025

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-----------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------|
| N1775-01 | DBRU7           | 1      | 1.17            | 8.41          | 9.58                     | 6.81                       | 67.1    |          |
| N1775-02 | DBRU8           | 2      | 1.11            | 8.69          | 9.8                      | 7.3                        | 71.2    |          |
| N1775-03 | DBRU9           | 3      | 1.11            | 8.63          | 9.74                     | 7.99                       | 79.7    |          |
| N1775-04 | DBRV0           | 4      | 1.12            | 8.75          | 9.87                     | 9.14                       | 91.7    |          |
| N1775-05 | DBRV1           | 5      | 1.14            | 8.72          | 9.86                     | 8.42                       | 83.5    |          |
| N1775-06 | DBRV2           | 6      | 1.13            | 8.51          | 9.64                     | 8.3                        | 84.3    |          |
| N1775-07 | DBRV3           | 7      | 1.19            | 8.79          | 9.98                     | 9.03                       | 89.2    |          |
| N1775-08 | DBRX0           | 8      | 1.19            | 8.75          | 9.94                     | 9.13                       | 90.7    |          |
| N1775-11 | VHBLK002        | 9      | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | VHBLK    |
| N1775-12 | DBRU6           | 10     | 1.15            | 8.81          | 9.96                     | 8.11                       | 79.0    |          |
| N1775-13 | DBRU6MS         | 11     | 1.15            | 8.81          | 9.96                     | 8.11                       | 79.0    |          |
| N1775-14 | DBRU6MSD        | 12     | 1.15            | 8.81          | 9.96                     | 8.11                       | 79.0    |          |
| N1775-15 | DBSB8           | 13     | 1.13            | 8.47          | 9.6                      | 8.46                       | 86.5    |          |
| N1775-16 | DBSB9           | 14     | 1.14            | 8.75          | 9.89                     | 8.85                       | 88.1    |          |
| N1775-17 | DBSC0           | 15     | 1.16            | 8.80          | 9.96                     | 8.68                       | 85.5    |          |
| N1775-18 | DBSC1           | 16     | 1.13            | 8.79          | 9.92                     | 7.04                       | 67.2    |          |
| N1775-19 | DBSC2           | 17     | 1.16            | 8.81          | 9.97                     | 7.77                       | 75.0    |          |
| N1775-20 | DBSC3           | 18     | 1.13            | 8.69          | 9.82                     | 7.68                       | 75.4    |          |
| N1775-21 | DBSC4           | 19     | 1.12            | 8.50          | 9.62                     | 7.38                       | 73.6    |          |
| N1775-22 | DBSC5           | 20     | 1.16            | 8.49          | 9.65                     | 7.07                       | 69.6    |          |
| N1775-23 | DBSC6           | 21     | 1.13            | 8.63          | 9.76                     | 6.89                       | 66.7    |          |
| N1775-24 | DBSC7           | 22     | 1.13            | 8.84          | 9.97                     | 7.66                       | 73.9    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$