


LB119026

7471B

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.0591e-004

C= 2.4013e-002

Rho= 0.9998538

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.024	0.024	2	0.000	2					
0.2	0.200	0.191	-0.009	813	0.0 %	813					-5
2.5	2.500	2.492	-0.008	11987	0.0 %	11987					0
5.0	5.000	5.047	0.047	24396	0.0 %	24396					1
7.5	7.500	7.374	-0.126	35694	0.0 %	35694					-2
10.0	10.000	10.071	0.071	48791	0.0 %	48791					1

LB119026 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Mar 2022 13:37:35

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	2	-	0.00007471B	PPB	07 Mar 2022 13:44:07	S	Std
0.2 - 1	50.2	813	-	0.20007471B	PPB	07 Mar 2022 13:46:25	S	Std
2.5 - 1	52.5	11987	-	2.50007471B	PPB	07 Mar 2022 13:48:42	S	Std
5.0 - 1	55	24396	-	5.00007471B	PPB	07 Mar 2022 13:50:59	S	Std
7.5 - 1	57.5	35694	-	7.50007471B	PPB	07 Mar 2022 13:53:17	S	Std
10.0 - 1	51.0	48791	-	10.00007471B	PPB	07 Mar 2022 14:05:43	S	Std
ICV22 - 1	ICV22	20176	4.1785	-7471B	PPB	07 Mar 2022 14:08:25	U	SMPL
ICB22 - 1	ICB22	-35	0.0168	-7471B	PPB	07 Mar 2022 14:10:40	U	SMPL
CCV13 - 1	CCV13	25963	5.3701	-7471B	PPB	07 Mar 2022 14:12:56	U	SMPL
CCB13 - 1	CCB13	-47	0.0143	-7471B	PPB	07 Mar 2022 14:15:11	U	SMPL
CRA - 1	CRA	806	0.1900	-7471B	PPB	07 Mar 2022 14:17:26	U	SMPL
HighStd - 1	HighStd	50518	10.4262	-7471B	PPB	07 Mar 2022 14:19:42	U	SMPL
ChkStd - 1	ChkStd	36512	7.5423	-7471B	PPB	07 Mar 2022 14:21:57	U	SMPL
PB143149BL - 1	PBS	-75	0.0086	-7471B	PPB	07 Mar 2022 14:24:14	U	SMPL
PB143149BS - 1	LCSS	20514	4.2481	-7471B	PPB	07 Mar 2022 14:28:36	U	SMPL
N1802-02 - 1	342	1948	0.4251	-7471B	PPB	07 Mar 2022 14:30:52	U	SMPL
N1805-01 - 1	NB-08-030422	9328	1.9448	-7471B	PPB	07 Mar 2022 14:33:10	U	SMPL
N1806-01 - 1	DUM-22	1441	0.3207	-7471B	PPB	07 Mar 2022 14:35:28	U	SMPL
N1806-03 - 1	DUM-23	1353	0.3026	-7471B	PPB	07 Mar 2022 14:37:45	U	SMPL
N1806-05 - 1	DUM-24	1085	0.2474	-7471B	PPB	07 Mar 2022 14:40:00	U	SMPL
CCV14 - 1	CCV14	25426	5.2595	-7471B	PPB	07 Mar 2022 14:44:50	U	SMPL
CCB14 - 1	CCB14	-67	0.0102	-7471B	PPB	07 Mar 2022 14:47:05	U	SMPL
N1806-07 - 1	DUM-25	1584	0.3502	-7471B	PPB	07 Mar 2022 14:49:21	U	SMPL
N1806-09 - 1	DUM-26	691	0.1663	-7471B	PPB	07 Mar 2022 14:51:37	U	SMPL
N1806-11 - 1	DUM-27	2054	0.4470	-7471B	PPB	07 Mar 2022 14:53:52	U	SMPL
N1808-01 - 1	CL-01-030422	2758	0.5919	-7471B	PPB	07 Mar 2022 14:56:08	U	SMPL
N1809-01 - 1	OR-02-030422	1716	0.3774	-7471B	PPB	07 Mar 2022 14:58:25	U	SMPL
N1809-01DUP - 1	OR-02-030422DUP	1802	0.3951	-7471B	PPB	07 Mar 2022 15:00:41	U	SMPL
N1809-01MS - 1	OR-02-030422MS	21881	4.5296	-7471B	PPB	07 Mar 2022 15:05:32	U	SMPL
N1809-01MSD - 1	OR-02-030422MSD	21658	4.4836	-7471B	PPB	07 Mar 2022 15:07:48	U	SMPL
N1809-01LX5 - 1		288	0.0833	-7471B	PPB	07 Mar 2022 15:10:05	U	SMPL
N1809-01A - 1		22185	4.5922	-7471B	PPB	07 Mar 2022 15:12:21	U	SMPL
CCV15 - 1	CCV15	25501	5.2750	-7471B	PPB	07 Mar 2022 15:14:37	U	SMPL
CCB15 - 1	CCB15	-50	0.0137	-7471B	PPB	07 Mar 2022 15:16:53	U	SMPL