



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 3/9/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 16:45

In Date: 03/08/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:25

Out Date: 03/09/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB119034

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1825-01	DUM-28	1	1.14	8.51	9.65	8.51	86.6	
N1825-02	DUM-28-EPH-2	2	1.18	8.45	9.63	8.52	86.9	
N1825-03	DUM-29	3	1.18	8.60	9.78	8.64	86.7	
N1825-04	DUM-29-EPH-2	4	1.17	8.62	9.79	8.66	86.9	
N1825-05	DUM-30	5	1.15	8.80	9.95	9.13	90.7	
N1825-06	DUM-30-EPH-2	6	1.18	8.37	9.55	8.45	86.9	
N1825-07	DUM-31	7	1.14	8.83	9.97	8.89	87.8	
N1825-08	DUM-31-EPH-2	8	1.19	8.59	9.78	8.34	83.2	
N1825-09	DUM-33	9	1.14	8.84	9.98	8.7	85.5	
N1825-10	DUM-33-EPH-2	10	1.19	8.47	9.66	8.47	86.0	
N1825-11	DUM-32	11	1.16	8.50	9.66	8.65	88.1	
N1825-12	DUM-32-EPH-2	12	1.19	8.41	9.6	8.61	88.2	
N1827-01	CLI-2022-23A	13	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-02	CLI-2022-23B	14	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-03	CLI-2022-24A	15	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-04	CLI-2022-25A	16	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-05	CLI-2022-26A	17	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-06	CLI-2022-27A	18	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-07	CLI-2022-28A	19	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-08	CLI-2022-28B	20	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-09	CLI-2022-29A	21	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-10	CLI-2022-29B	22	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-11	HED953U-END-1-1	23	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-12	HED953U-END-1-2	24	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1827-13	RO-291	25	1.00	1.00	2.00	2.00	100.0	debris
N1827-14	RO-291-EPH-2	26	1.00	1.00	2.00	2.00	100.0	debris
N1828-01	BC260592-END-1-2	27	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1828-02	BC107651LOK-END-1-1	28	1.00	1.00	2.00	2.00	100.0	PILC REEL



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 3/9/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 16:45

In Date: 03/08/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:25

Out Date: 03/09/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB119034

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1828-03	BC107651LOK-END-1-2	29	1.00	1.00	2.00	2.00	100.0	PILC REEL
N1829-01	SU-03-030722	30	1.12	8.46	9.58	8.72	89.8	
N1829-02	SU-03-030722-EPH-2	31	1.15	8.80	9.95	9.34	93.1	
N1829-03	SU-04-030722	32	1.18	8.47	9.65	9.19	94.6	
N1829-04	SU-04-030722-EPH-2	33	1.17	8.58	9.75	9.05	91.8	
N1830-01	11986	34	1.14	8.77	9.91	8.96	89.2	
N1830-02	11986	35	1.14	8.77	9.91	8.96	89.2	
N1830-03	11986-EPH-2	36	1.19	8.63	9.82	8.91	89.5	
N1830-04	11977	37	1.11	8.55	9.66	9.11	93.6	
N1830-05	11977	38	1.11	8.55	9.66	9.11	93.6	
N1830-06	11977-EPH-2	39	1.18	8.57	9.75	8.82	89.1	
N1830-07	350	40	1.18	8.57	9.75	8.82	89.1	
N1830-08	350	41	1.18	8.57	9.75	8.82	89.1	
N1830-09	350-EPH-2	42	1.18	8.76	9.94	9.29	92.6	
N1831-01	AU-06-03072022	43	1.14	8.58	9.72	8.25	82.9	
N1832-01	BU-02-03072022	44	1.15	8.81	9.96	9.72	97.3	
N1833-01	TR-05-03072022	45	1.18	8.64	9.82	9.05	91.1	
N1833-02	TR-05-03072022-EPH-2	46	1.15	8.81	9.96	9.3	92.5	
N1834-01	AUD-1281	47	1.11	8.59	9.7	9.4	96.5	
N1835-04	AUD-1272	48	1.17	8.54	9.71	8.34	84.0	
N1835-07	AUD-1242	49	1.00	1.00	2.00	2.00	100.0	debris
N1843-01	FS-1	50	1.18	8.49	9.67	8.48	86.0	
N1843-02	FS-1	51	1.15	8.50	9.65	8.56	87.2	
N1843-03	FS-2	52	1.17	8.49	9.66	8.57	87.2	
N1843-04	FS-2	53	1.11	8.56	9.67	8.4	85.2	
N1845-01	SAMPLE-1	54	1.14	8.74	9.88	9.05	90.5	
N1845-02	SAMPLE-2	55	1.18	8.45	9.63	8.7	89.0	
N1845-03	SAMPLE-3	56	1.18	8.52	9.7	8.85	90.0	



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 3/9/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 16:45

In Date: 03/08/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:25

Out Date: 03/09/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB119034

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1845-04	SAMPLE-4	57	1.15	8.40	9.55	8.79	91.0	
N1845-05	SAMPLE-5	58	1.17	8.77	9.94	9.14	90.9	
N1845-06	SAMPLE-6	59	1.18	8.52	9.7	8.68	88.0	
N1845-07	SAMPLE-7	60	1.14	8.81	9.95	9.28	92.4	
N1845-08	SAMPLE-8	61	1.14	8.81	9.95	9.28	92.4	
N1845-09	SAMPLE-9	62	1.15	8.51	9.66	8.7	88.7	
N1845-10	SAMPLE-10	63	1.19	8.36	9.55	8.68	89.6	
N1845-11	SAMPLE-11	64	1.15	8.61	9.76	8.74	88.2	
N1845-12	SAMPLE-12	65	1.17	8.60	9.77	8.87	89.5	
N1845-13	SAMPLE-13	66	1.11	8.72	9.83	8.88	89.1	
N1845-14	SAMPLE-14	67	1.18	8.48	9.66	8.67	88.3	
N1845-15	SAMPLE-15	68	1.15	8.45	9.6	8.72	89.6	
N1851-01	DUM-34	69	1.14	8.53	9.67	8.64	87.9	
N1851-02	DUM-34-EPH-2	70	1.14	8.67	9.81	8.48	84.7	
N1851-03	DUM-35	71	1.14	8.80	9.94	8.97	89.0	
N1851-04	DUM-35-EPH-2	72	1.19	8.61	9.8	8.77	88.0	
N1851-05	DUM-36	73	1.19	8.77	9.96	9.00	89.1	
N1851-06	DUM-36-EPH-2	74	1.15	8.80	9.95	9.02	89.4	
N1851-07	DUM-37	75	1.13	8.66	9.79	8.94	90.2	
N1851-08	DUM-37-EPH-2	76	1.19	8.52	9.71	8.69	88.0	
N1851-09	DUM-38	77	1.13	8.50	9.63	8.88	91.2	
N1851-10	DUM-38-EPH-2	78	1.18	8.44	9.62	8.79	90.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$