


LB119036

SFAM01.1

INSTRUMENT 111. CV2

 μ Abs.:

41531



Linear

A= 0.0000e+000

B= 2.4040e-004

C= 6.6565e-002

Rho= 0.9998788

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.070	0.070	16	0.000	16					=
0.05	0.050					185	0				14
0.20	0.200	0.228	0.028	672	0.0 %	672					-3
2.50	2.500	2.416	-0.084	9772	0.0 %	9772					-1
5.00	5.000	4.934	-0.066	20246	0.0 %	20246					0
7.50	7.500	7.501	0.001	30927	0.0 %	30927					1
10.0	10.000	10.051	0.051	41531	0.0 %	41531					

LB119036
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	16	-		0 SFAM01.1	PPB	3/8/2022 10:51	Std
	0.2 S01	672	-		0.2 SFAM01.1	PPB	3/8/2022 10:53	Std
	2.5 S02	9772	-		2.5 SFAM01.1	PPB	3/8/2022 10:55	Std
	5 S03	20246	-		5 SFAM01.1	PPB	3/8/2022 10:57	Std
	7.5 S04	30927	-		7.5 SFAM01.1	PPB	3/8/2022 11:00	Std
	10 S05	41531	-		10 SFAM01.1	PPB	3/8/2022 11:02	Std
ICV023	ICV023	15661	3.8315 -		SFAM01.1	PPB	3/8/2022 11:05	SMPL
ICB023	ICB023	-22	0.0613 -		SFAM01.1	PPB	3/8/2022 11:07	SMPL
CCV016	CCV016	20205	4.9238 -		SFAM01.1	PPB	3/8/2022 11:09	SMPL
CCB016	CCB016	-16	0.0627 -		SFAM01.1	PPB	3/8/2022 11:12	SMPL
PB143176BL	PBW176	84	0.0868 -		SFAM01.1	PPB	3/8/2022 11:14	SMPL
N1737-01	YAY93	236590	56.9428 -		SFAM01.1	PPB	3/8/2022 11:16	SMPL
N1737-02	YAY99	7318	1.8258 -		SFAM01.1	PPB	3/8/2022 11:18	SMPL
N1737-03	YAYA1	160478	38.6454 -		SFAM01.1	PPB	3/8/2022 11:21	SMPL
N1737-04	YAYA1D	159461	38.401 -		SFAM01.1	PPB	3/8/2022 11:23	SMPL
N1737-05	YAYA1S	160566	38.6666 -		SFAM01.1	PPB	3/8/2022 11:25	SMPL
N1737-06	YAYA7	3511	0.9106 -		SFAM01.1	PPB	3/8/2022 11:28	SMPL
N1737-07	YAYE7	6510	1.6316 -		SFAM01.1	PPB	3/8/2022 11:30	SMPL
N1737-08	YAYB9	112	0.0935 -		SFAM01.1	PPB	3/8/2022 11:32	SMPL
N1737-09	YAYC9	94521	22.7894 -		SFAM01.1	PPB	3/8/2022 11:35	SMPL
N1737-10	YAYD3	199767	48.0905 -		SFAM01.1	PPB	3/8/2022 11:37	SMPL
N1737-11	YAYD9	-276	0.0002 -		SFAM01.1	PPB	3/8/2022 11:39	SMPL
N1737-12	YAYD5	38985	9.4386 -		SFAM01.1	PPB	3/8/2022 11:41	SMPL
N1737-13	YAYF6	-96	0.0435 -		SFAM01.1	PPB	3/8/2022 11:44	SMPL
N1737-14	YAYF7	-26	0.0603 -		SFAM01.1	PPB	3/8/2022 11:46	SMPL
PB143177BL	PBW177	-10	0.0642 -		SFAM01.1	PPB	3/8/2022 11:48	SMPL
N1738-01	YAY94	68978	16.6489 -		SFAM01.1	PPB	3/8/2022 11:51	SMPL
N1738-02	YAYA0	696	0.2339 -		SFAM01.1	PPB	3/8/2022 11:53	SMPL
N1738-03	YAYA2	178565	42.9936 -		SFAM01.1	PPB	3/8/2022 11:55	SMPL
N1738-04	YAYA2D	184680	44.4636 -		SFAM01.1	PPB	3/8/2022 11:57	SMPL
CCV017	CCV017	19713	4.8056 -		SFAM01.1	PPB	3/8/2022 12:00	SMPL
CCB017	CCB017	-85	0.0461 -		SFAM01.1	PPB	3/8/2022 12:02	SMPL
N1738-05	YAYA2S	186313	44.8562 -		SFAM01.1	PPB	3/8/2022 12:04	SMPL
N1738-06	YAYA8	74	0.0844 -		SFAM01.1	PPB	3/8/2022 12:07	SMPL
N1738-07	YAYE8	548	0.1983 -		SFAM01.1	PPB	3/8/2022 12:09	SMPL
N1738-08	YAYC0	48	0.0781 -		SFAM01.1	PPB	3/8/2022 12:11	SMPL
N1738-09	YAYD0	93335	22.5043 -		SFAM01.1	PPB	3/8/2022 12:13	SMPL
N1738-10	YAYD4	201828	48.586 -		SFAM01.1	PPB	3/8/2022 12:16	SMPL
N1738-11	YAYE0	-293	-0.0039 -		SFAM01.1	PPB	3/8/2022 12:18	SMPL
N1738-12	YAYD6	584	0.207 -		SFAM01.1	PPB	3/8/2022 12:20	SMPL
N1738-13	YAYF5	-21	0.0615 -		SFAM01.1	PPB	3/8/2022 12:23	SMPL
PB143178BL	PBW178	77	0.0851 -		SFAM01.1	PPB	3/8/2022 12:25	SMPL
N1741-01	YAY95	96608	23.2911 -		SFAM01.1	PPB	3/8/2022 12:27	SMPL
N1741-02	YAY95D	96266	23.2089 -		SFAM01.1	PPB	3/8/2022 12:29	SMPL
N1741-03	YAY95S	98319	23.7024 -		SFAM01.1	PPB	3/8/2022 12:32	SMPL
PB143179BL	PBW179	-55	0.0533 -		SFAM01.1	PPB	3/8/2022 12:34	SMPL

N1743-01	YAY96	100486	24.2234 -	SFAM01.1	PPB	3/8/2022 12:36	SMPL
N1743-02	YAY96D	101191	24.3929 -	SFAM01.1	PPB	3/8/2022 12:38	SMPL
N1743-03	YAY96S	104368	25.1566 -	SFAM01.1	PPB	3/8/2022 12:41	SMPL
PB143180BL	PBW180	-66	0.0507 -	SFAM01.1	PPB	3/8/2022 12:43	SMPL
N1784-01	DBRU0	18	0.0709 -	SFAM01.1	PPB	3/8/2022 12:45	SMPL
N1784-02	DBRU1	25	0.0726 -	SFAM01.1	PPB	3/8/2022 12:48	SMPL
CCV018	CCV018	20653	5.0315 -	SFAM01.1	PPB	3/8/2022 12:50	SMPL
CCB018	CCB018	-43	0.0562 -	SFAM01.1	PPB	3/8/2022 12:52	SMPL
N1737-01DLX10	YAY93	27460	6.6679 -	SFAM01.1	PPB	3/8/2022 12:54	SMPL
N1737-03DLX10	YAYA1	17712	4.3245 -	SFAM01.1	PPB	3/8/2022 12:57	SMPL
N1737-04DLX10	YAYA1D	18406	4.4914 -	SFAM01.1	PPB	3/8/2022 12:59	SMPL
N1737-05DLX10	YAYA1S	18446	4.501 -	SFAM01.1	PPB	3/8/2022 13:01	SMPL
N1737-09DLX5	YAYC9	20891	5.0888 -	SFAM01.1	PPB	3/8/2022 13:04	SMPL
N1737-10DLX10	YAYD3	25293	6.147 -	SFAM01.1	PPB	3/8/2022 13:06	SMPL
N1738-01DLX5	YAY94	15577	3.8113 -	SFAM01.1	PPB	3/8/2022 13:08	SMPL
N1738-03DLX10	YAYA2	20334	4.9549 -	SFAM01.1	PPB	3/8/2022 13:10	SMPL
N1738-04DLX10	YAYA2D	19828	4.8332 -	SFAM01.1	PPB	3/8/2022 13:13	SMPL
N1738-05DLX10	YAYA2S	20560	5.0092 -	SFAM01.1	PPB	3/8/2022 13:15	SMPL
N1738-09DLX5	YAYD0	19688	4.7996 -	SFAM01.1	PPB	3/8/2022 13:17	SMPL
N1738-10DLX10	YAYD4	25345	6.1595 -	SFAM01.1	PPB	3/8/2022 13:19	SMPL
CCV019	CCV019	20446	4.9818 -	SFAM01.1	PPB	3/8/2022 13:22	SMPL
CCB019	CCB019	-27	0.0601 -	SFAM01.1	PPB	3/8/2022 13:24	SMPL
N1741-01DLX5	YAY95	18444	4.5005 -	SFAM01.1	PPB	3/8/2022 13:26	SMPL
N1741-02DLX5	YAY95D	19182	4.6779 -	SFAM01.1	PPB	3/8/2022 13:29	SMPL
N1741-03DLX5	YAY95S	19999	4.8743 -	SFAM01.1	PPB	3/8/2022 13:31	SMPL
N1743-01DLX5	YAY96	20966	5.1068 -	SFAM01.1	PPB	3/8/2022 13:33	SMPL
N1743-02DLX5	YAY96D	20944	5.1015 -	SFAM01.1	PPB	3/8/2022 13:35	SMPL
N1743-03DLX5	YAY96S	22533	5.4835 -	SFAM01.1	PPB	3/8/2022 13:38	SMPL
CCV020	CCV020	20191	4.9205 -	SFAM01.1	PPB	3/8/2022 13:42	SMPL
CCB020	CCB020	-15	0.063 -	SFAM01.1	PPB	3/8/2022 13:45	SMPL