


LB119048

7471B

INSTRUMENT ID: CV2

 μ Abs.:

49667



Linear

A= 0.0000e+000

B= 1.9419e-004

C= 1.0242e-002

Rho= 0.9980479

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/61
0.0	0.000	0.013	0.013	16	0.000	16					
0.2	0.200	0.187	-0.013	912	0.0 %	912					-7
2.5	2.500	2.339	-0.161	11992	0.0 %	11992					-6
5.0	5.000	5.101	0.101	26218	0.0 %	26218					2
7.5	7.500	7.904	0.404	40649	0.0 %	40649					5
10.0	10.000	9.655	-0.345	49667	0.0 %	49667					-3

LB119048 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Mar 2022 14:42:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	16	-	0.0000	7471B	PPB	08 Mar 2022 14:48:00	S	Std
0.2 - 1	50.2	912	-	0.2000	7471B	PPB	08 Mar 2022 14:50:18	S	Std
2.5 - 1	52.5	11992	-	2.5000	7471B	PPB	08 Mar 2022 14:52:36	S	Std
5.0 - 1	55	26218	-	5.0000	7471B	PPB	08 Mar 2022 14:54:53	S	Std
7.5 - 1	57.5	40649	-	7.5000	7471B	PPB	08 Mar 2022 14:57:10	S	Std
10.0 - 1	610	49667	-	10.0000	7471B	PPB	08 Mar 2022 14:59:26	S	Std
ICV24 - 1	ICV24	20218	3.9364	-	7471B	PPB	08 Mar 2022 15:02:27	U	SMPL
ICB24 - 1	ICB24	-14	0.0075	-	7471B	PPB	08 Mar 2022 15:04:42	U	SMPL
CCV21 - 1	CCV21	25354	4.9337	-	7471B	PPB	08 Mar 2022 15:17:11	U	SMPL
CCB21 - 1	CCB21	-43	0.0019	-	7471B	PPB	08 Mar 2022 15:19:27	U	SMPL
CRA - 1	CRA	790	0.1637	-	7471B	PPB	08 Mar 2022 15:21:43	U	SMPL
HighStd - 1	HighStd	46640	9.0672	-	7471B	PPB	08 Mar 2022 15:26:32	U	SMPL
ChkStd - 1	ChkStd	32574	6.3358	-	7471B	PPB	08 Mar 2022 15:28:47	U	SMPL
PB143186BL - 1	PBS	-71	-0.0035	-	7471B	PPB	08 Mar 2022 15:31:05	U	SMPL
PB143186BS - 1	LCSS	18191	3.5427	-	7471B	PPB	08 Mar 2022 15:33:22	U	SMPL
N1825-01 - 1	DUM-28	3352	0.6612	-	7471B	PPB	08 Mar 2022 15:35:39	U	SMPL
N1825-03 - 1	DUM-29	2112	0.4204	-	7471B	PPB	08 Mar 2022 15:37:56	U	SMPL
N1825-05 - 1	DUM-30	8164	1.5956	-	7471B	PPB	08 Mar 2022 15:40:13	U	SMPL
N1825-07 - 1	DUM-31	2534	0.5023	-	7471B	PPB	08 Mar 2022 15:42:30	U	SMPL
N1825-09 - 1	DUM-33	2717	0.5379	-	7471B	PPB	08 Mar 2022 15:44:46	U	SMPL
CCV22 - 1	CCV22	27095	5.2718	-	7471B	PPB	08 Mar 2022 15:47:02	U	SMPL
CCB22 - 1	CCB22	-27	0.0050	-	7471B	PPB	08 Mar 2022 15:49:18	U	SMPL
N1825-11 - 1	DUM-32	2821	0.5580	-	7471B	PPB	08 Mar 2022 15:51:34	U	SMPL
N1825-11DUP - 1	DUM-32DUP	3252	0.6417	-	7471B	PPB	08 Mar 2022 15:53:50	U	SMPL
N1825-11MS - 1	DUM-32MS	22270	4.3348	-	7471B	PPB	08 Mar 2022 15:56:06	U	SMPL
N1825-11MSD - 1	DUM-32MSD	21792	4.2420	-	7471B	PPB	08 Mar 2022 15:58:22	U	SMPL
N1827-13 - 1	RO-291	105	0.0306	-	7471B	PPB	08 Mar 2022 16:00:38	U	SMPL
N1829-01 - 1	SU-03-030722	1729	0.3460	-	7471B	PPB	08 Mar 2022 16:02:55	U	SMPL
N1829-03 - 1	SU-04-030722	857	0.1767	-	7471B	PPB	08 Mar 2022 16:05:11	U	SMPL
N1830-01 - 1	11986	1953	0.3895	-	7471B	PPB	08 Mar 2022 16:07:28	U	SMPL
N1830-04 - 1	11977	6247	1.2233	-	7471B	PPB	08 Mar 2022 16:09:46	U	SMPL
N1830-07 - 1	350	5744	1.1257	-	7471B	PPB	08 Mar 2022 16:12:01	U	SMPL
CCV23 - 1	CCV23	26076	5.0739	-	7471B	PPB	08 Mar 2022 16:14:17	U	SMPL
CCB23 - 1	CCB23	-25	0.0054	-	7471B	PPB	08 Mar 2022 16:16:33	U	SMPL
N1831-01 - 1	AU-06-03072022	8722	1.7040	-	7471B	PPB	08 Mar 2022 16:18:48	U	SMPL
N1832-01 - 1	BU-02-03072022	582	0.1233	-	7471B	PPB	08 Mar 2022 16:21:04	U	SMPL
N1833-01 - 1	TR-05-03072022	13278	2.5887	-	7471B	PPB	08 Mar 2022 16:23:21	U	SMPL
N1825-11LX5 - 1		528	0.1128	-	7471B	PPB	08 Mar 2022 16:25:37	U	SMPL
N1825-11A - 1		23894	4.6502	-	7471B	PPB	08 Mar 2022 16:27:53	U	SMPL
CCV24 - 1	CCV24	26676	5.1904	-	7471B	PPB	08 Mar 2022 16:30:10	U	SMPL
CCB24 - 1	CCB24	-19	0.0066	-	7471B	PPB	08 Mar 2022 16:32:27	U	SMPL